



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 8:56 am, Oct 11, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102469	PRINTER SN 095.353.193	DATE OF INSPECTION 10/08/2022
-----------------------------	---------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) Lafayette County Sheriff 107 S. 11th Lexington, Mo. 64067	TIME OF INSPECTION 2:05 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 21380 EXP. DATE 09/13/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIMULATOR SN SD2275 SIMULATOR EXP DATE 01/18/2023

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .103	TEST 2 ➡ .103	TEST 3 ➡ .099
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Meets DOH Standards.

INSPECTING OFFICER

SIGNATURE #156	PRINT NAME Caleb M. C. Crosson
-------------------	-----------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 220164 06/24/2024	TELEPHONE NUMBER (660) 259-3622
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 102469
Version no: 582B

TEST RECORD 00829

Temp Date Time 210L
g/

Air Blank: 10/08/22 04:52 .009
Calibration Check: 24 10/08/22 04:52 .103

Subject Name

CAL CHECK

Subject I.D.

Operator Name, I.D.

C. Casson 220164

Location

LC50

AS IV Serial no: 102469
Version no: 582B

TEST RECORD 00830

Temp Date Time 210L
g/

Air Blank: 10/08/22 04:54 .003
Calibration Check: 24 10/08/22 04:54 .103

Subject Name

CAL CHECK

Subject I.D.

Operator Name, I.D.

C. Casson 220164

Location

LC50

AS IV Serial no: 102469
Version no: 582B

TEST RECORD 00831

Temp Date Time 210L
g/

Air Blank: 10/08/22 04:55 .000
Calibration Check: 25 10/08/22 04:56 .099

Subject Name

CAL CHECK

Subject I.D.

Operator Name, I.D.

C. Casson 220164

Location

LC50

AS IV Serial no: 102469
Version no: 582B

TEST RECORD 00832

Temp Date Time 210L
g/

VOID: 221
12 10/08/22 04:57

Subject Name

CAL CHECK

Subject I.D.

Operator Name, I.D.

C. Casson 220164

Location

LC50



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21380** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 15, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 13, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
CALEB CROSSON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

NUMBER 220164

EXPIRES 6/24/2024

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula F. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator CROSSON, CALEB
 Permit No 220164
 Date Issued 6/24/2022 Date Expires 6/24/2024

