



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 7:40 am, Jun 09, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102473	NAME OF AGENCY Greene County Sheriff's Office	DATE OF INSPECTION 04/22/2022
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 1010 North Boonville Avenue Springfield, Missouri	TIME OF INSPECTION 3:01 pm
---	-------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u>	LOT # <u>AG111803</u> EXP. DATE <u>04/28/2023</u>
<input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C)	SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .097	TEST 2  .096	TEST 3  .096
--------------	--------------	--------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Kyle Winchell
TYPE II PERMIT NUMBER/EXPIRATION DATE 200259 09/24/2022	TELEPHONE NUMBER (417) 868-4040

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**KYLE R. WINCHELL**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/24/2020

NUMBER 200259

EXPIRES 9/24/2022

MO 580-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator WINCHELL, KYLE  
 Permit No 200259  
 Date Issued 9/24/2020 Date Expires 9/24/2022





Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

**Test Date:** 28-Apr-2021

**Lot # AG111803 Model 108cacc**

**Exp. Date**

28-Apr-2023

**Cyl. Type**

108

**Component**

Ethanol

Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)

Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

**RGM Serial No.**

EB0010581

**Concentration**

392.1 ppm

EB0010570

259.8 ppm

EB0010285

208.0 ppm

EB0010561

103.6 ppm

EB0010681

52.12 ppm

**RGM Serial No.**

EB0010603

**Concentration**

393.0 ppm

EB0010559

258.2 ppm

EB0010595

208.3 ppm

EB0010562

104.2 ppm

EB0010579

52.81 ppm

**CRM Serial No.**

CC434668

**Concentration**

800.0 ppm

CC234503

253.0 ppm

**CRM Serial No.**

0056649

**Concentration**

390.1 ppm

0056662

150.2 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2021.04.28 18:40:40 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**

**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 01773

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
04/22/22 15:01 .000  
Calibration Check:  
24 04/22/22 15:01 .097

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

ZL Zwell

Location

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 01774

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
04/22/22 15:03 .000  
Calibration Check:  
25 04/22/22 15:03 .096

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

ZL Zwell

Location

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 01775

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
04/22/22 15:04 .000  
Calibration Check:  
25 04/22/22 15:04 .096

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

ZL Zwell

Location

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 01776

Temp Date Time <sup>a/</sup> 210L

VOID: RFI  
12 04/22/22 15:05

Subject Name

Test 4 / RFI

Subject I.D.

Operator Name, I.D.

ZL Zwell

Location

AS IV Serial no: 102473  
Version no: 532B

TEST RECORD 01777

Temp Date Time <sup>a/</sup> 210L

Air Blank:  
04/22/22 15:07 .000  
Subject Test: Auto  
26 04/22/22 15:07 .000

Subject Name

Test 5 / Sobu

Subject I.D.

Operator Name, I.D.

ZL Zwell

Location