

RECEIVED

By Tracy Crews at 7:47 am, Dec 27, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 105445	NAME OF AGENCY Platte County Sheriff's Office	DATE OF INSPECTION 12/22/2022
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 415 Third Street, Platte City, Missouri 64079	TIME OF INSPECTION 7:40 pm
---	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters, Inc. LOT # AG114002 EXP. DATE 05/20/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .102	TEST 2 • .101	TEST 3 • .101
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04) 1	(.05-.09) 1	(.10-.14)	(.15-.19)	(OVER .19)
----------	-----------	-------------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>William Beeler</i>	PRINT NAME William Beeler
------------------------------------	------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 220122 / 05-11-2024	TELEPHONE NUMBER (816) 858-3521
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01171

Temp Date Time ^{g/} 210L

Air Blank:
12/22/22 19:40 .000
Calibration Check:
18 12/22/22 19:40 .102

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

W. Beeler ²²⁰¹²² 5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01172

Temp Date Time ^{g/} 210L

Air Blank:
12/22/22 19:42 .000
Calibration Check:
19 12/22/22 19:42 .101

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

W. Beeler ²²⁰¹²² 5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01173

Temp Date Time ^{g/} 210L

VOID: BAT LOW
1 12/22/22 19:44

Subject Name

Battery Low

Subject I.D.

Operator Name, I.D.

W. Beeler ²²⁰¹²² 5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01174

Temp Date Time ^{g/} 210L

Air Blank:
12/22/22 19:46 .000
Calibration Check:
21 12/22/22 19:46 .101

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

W. Beeler ²²⁰¹²² 5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01175

Temp Date Time ^{g/} 210L

VOID: RFI
12 12/22/22 19:48

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

W. Beeler ²²⁰¹²² 5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 105445
Version no: 532B

TEST RECORD 01176

Temp Date Time ^{g/} 210L

Air Blank:
12/22/22 19:49 .000
Subject Test: Auto
22 12/22/22 19:49 .000

Subject Name

Blank

Subject I.D.

Operator Name, I.D.

W. Beeler ²²⁰¹²² 5/11/2024

Location

415 Third St

Platte City MO



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 21-May-2021

Lot # AG114002 Model 108cacc

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
20-May-2023	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2021.05.21 11:39:32 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
WILLIAM BEELER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

NUMBER 220122

EXPIRES 5/11/2024

Laura P. Wang

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BEELER, WILLIAM
Permit No 220122
Date Issued 5/11/2022 **Date Expires** 5/11/2024

