

RECEIVED

By Tracy Crews at 3:54 pm, Nov 21, 2022

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107972	NAME OF AGENCY BOONE COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 11/20/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 2111 E. COUNTY DR, COLUMBIA		TIME OF INSPECTION 8:35 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER INTOXIMETERS LOT # AG112301 EXP. DATE 05/03/2023 SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100

TEST 2 .100

TEST 3 .100

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	1	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT OPERATING SATISFACTORILY AND WITHIN ESTABLISHED LIMITS. CLOCK ADJUSTED FOR DAYLIGHT SAVING TIME.

INSPECTING OFFICER

SIGNATURE

PRINT NAME

STEVEN H. VERBLE

TYPE II PERMIT NUMBER/EXPIRATION DATE

210084, 04/22/2023

TELEPHONE NUMBER

(573) 875-1111

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 107972
Version no: 502B

TEST RECORD 00810

Temp Date Time 210L %/
Air Blank: 11/20/22 08:42 .000
Calibration Check: 22 11/20/22 08:42 .100

Subject Name
Test 1
Subject I.D.

Operator Name, I.D.
Verble, 210084
Location
2111 E. County Dr,
Columbia

AS IV Serial no: 107972
Version no: 502B

TEST RECORD 00814

Temp Date Time 210L %/
Air Blank: 11/20/22 08:48 .000
Calibration Check: 24 11/20/22 08:48 .100

Subject Name
Test 2
Subject I.D.

Operator Name, I.D.
Verble, 210084
Location
2111 E. County Dr
Columbia

AS IV Serial no: 107972
Version no: 502B

TEST RECORD - REPRINT

TEST RECORD 00815
Temp Date Time 210L %/
Air Blank: 11/20/22 09:55 .000
Calibration Check: 25 11/20/22 09:55 .100

Subject Name
Test 3
Subject I.D.

Operator Name, I.D.
Verble, 210084
Location
2111 E. County Dr-
Columbia

AS IV Serial no: 107972
Version no: 502B

TEST RECORD 00816

Temp Date Time 210L %/
UCID: RFI
12 11/20/22 09:00

Subject Name
RFI
Subject I.D.

Operator Name, I.D.
Verble, 210084
Location
2111 E. County Dr
Columbia



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 3-May-2021

Lot # AG112301 Model 108caod

Exp. Date 3-May-2023
Cyl. Type 108
Component Ethanol
Nitrogen
Certified Concentration 0.100 ± 2% BrAC (272 ppm)
Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010561	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010661	52.12 ppm	EB0010579	52.61 ppm
CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Printed: 05/03/2021 12:15:54 PM
Date: 2021-05-03 12:15:54 PM
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release: Paul Marsala
Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07

STEVEN H. VERBLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/22/2021
NUMBER 210084
EXPIRES 4/22/2023
MO 885-0771 (6-10)

Steven H. Verble
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paul Marsala
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB# (PH 10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named contributor is authorized to use the following instrument for the determination of the alcoholic content of breath from a sample of expired air in Missouri.

Operator: VERBLE STEVEN
Date Issued: 4/22/2021
Date Expires: 4/22/2023