

RECEIVED

By Tracy Crews at 7:54 am, Apr 18, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107973	PRINTER SN 099.3586.816	DATE OF INSPECTION 04/15/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 690 Chesterfield Pkwy W, Chesterfield MO		TIME OF INSPECTION 4:18 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG035102 EXP. DATE 12/16/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← 0.081

TEST 2 ← 0.081

TEST 3 ← 0.081

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 (0-.04) 0 (.05-.09) 0 (.10-.14) 0 (.15-.19) 0 (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within the Missouri Department of Health Specifications.

INSPECTING OFFICER

SIGNATURE

PRINT NAME

G. Rupp #264

TYPE II PERMIT NUMBER/EXPIRATION DATE

TELEPHONE NUMBER

636-537-3000

210082 / 04/21/2023

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 107973
Version no: 532B
TEST RECORD 00882 s/
Temp Date Time 210L
Air Blank: 04/15/22 16:21 .000
Calibration Check: 21 04/15/22 16:21 .081
Subject Name RFI Test
Subject I.D. G. Rupp # 264
Operator Name, I.D. 210082
Location 690 Chesterfield
Pkwy W

AS IV Serial no: 107973
Version no: 532B
TEST RECORD 00881 s/
Temp Date Time 210L
Air Blank: 04/15/22 16:21 .000
Calibration Check: 21 04/15/22 16:21 .081
Subject Name Test # 3
Subject I.D. G. Rupp # 264
Operator Name, I.D. 210082
Location 690 Chesterfield
Pkwy W

AS IV Serial no: 107973
Version no: 532B
TEST RECORD 00880 s/
Temp Date Time 210L
Air Blank: 04/15/22 16:19 .000
Calibration Check: 20 04/15/22 16:19 .081
Subject Name Test # 2
Subject I.D. G. Rupp # 264
Operator Name, I.D. 210082
Location 690 Chesterfield
Pkwy W

AS IV Serial no: 107973
Version no: 532B
TEST RECORD 00879 s/
Temp Date Time 210L
Air Blank: 04/15/22 16:18 .000
Calibration Check: 20 04/15/22 16:18 .081
Subject Name Test # 1
Subject I.D. G. Rupp # 264
Operator Name, I.D. 210082
Location 690 Chesterfield
Pkwy W



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 16-Dec-2020

Lot # AG035102 **Model** 108cacd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
16-Dec-2022	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (218 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2020.12.16 17:41:58 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
GREGORY RUPP

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/21/2021

NUMBER 210082

EXPIRES 4/21/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator RUPP, GREGORY
Permit No 210082
Date Issued 4/21/2021 **Date Expires** 4/21/2023

