

**RECEIVED**

By Tracy Crews at 9:32 pm, Feb 11, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107984	PRINTER SN 099.3586.817	DATE OF INSPECTION 02/04/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 111 East Court Street, Kahoka		TIME OF INSPECTION 10:57 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Repcos Marketing Inc. LOT # 20001 EXP. DATE 10/07/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIMULATOR SN MP2457 SIMULATOR EXP DATE 05/05/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .101

TEST 2 ➡ .100

TEST 3 ➡ .100

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Re-Calibrated at .100

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME  
Trooper Jonathan Conrad

TYPE II PERMIT NUMBER/EXPIRATION DATE  
200097/ Expires 02/18/2022

TELEPHONE NUMBER  
(660) 385-2132

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Boulevard  
Poplar Bluff, MO 63901

AS IV Serial no: 107984  
Version no: 532B

TEST RECORD 00651

Temp	Date	Time	a/ 210L
Air Blank:			
	02/04/22	22:57	.000
Calibration Check:			
	25 02/04/22	22:57	.101

Air Blank:  
02/04/22 22:57 .000  
Calibration Check:  
25 02/04/22 22:57 .101

Subject Name

TEST #1

Subject I.D.

Operator Name, I.D.

J. M. CONRAD

Location

TROOP B - ZONE 6 OFFICE

KAHOKA

AS IV Serial no: 107984  
Version no: 532B

TEST RECORD 00652

Temp	Date	Time	a/ 210L
Air Blank:			
	02/04/22	22:59	.000
Calibration Check:			
	26 02/04/22	22:59	.100

Air Blank:  
02/04/22 22:59 .000  
Calibration Check:  
26 02/04/22 22:59 .100

Subject Name

TEST #2

Subject I.D.

Operator Name, I.D.

J. M. CONRAD

Location

TROOP B - ZONE 6 OFFICE

KAHOKA

AS IV Serial no: 107984  
Version no: 532B

TEST RECORD 00653

Temp	Date	Time	a/ 210L
Air Blank:			
	02/04/22	23:01	.000
Calibration Check:			
	27 02/04/22	23:01	.100

Air Blank:  
02/04/22 23:01 .000  
Calibration Check:  
27 02/04/22 23:01 .100

Subject Name

TEST #3

Subject I.D.

Operator Name, I.D.

J. M. CONRAD

Location

TROOP B - ZONE 6 OFFICE

KAHOKA

AS IV Serial no: 107984  
Version no: 532B

TEST RECORD 00654

Temp	Date	Time	a/ 210L
VOID: RFI			
	12 02/04/22	23:03	

VOID: RFI  
12 02/04/22 23:03

Subject Name

RFI TEST

Subject I.D.

Operator Name, I.D.

J. M. CONRAD

Location

TROOP B - ZONE 6 OFFICE

KAHOKA



RepCo Marketing Co  
3101-188 Stony Brook Drive  
Raleigh, NC 27604  
919-876-5480

## CERTIFICATE OF ANALYSIS

**MANUFACTURER AND SUPPLIER: RepCo Marketing Co.**  
**LOT NUMBER: 20001**  
**EXPIRATION DATE: October 7, 2022 at 11:59 p.m.**

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

  
Alma Palmer, Operations Manager  
RepCo Marketing Co.



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**JONATHAN M CONRAD**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/18/2020

NUMBER 200097

EXPIRES 2/18/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator CONRAD, JONATHAN  
 Permit No 200097  
 Date Issued 2/18/2020 Date Expires 2/18/2022

