

RECEIVED

By Tracy Crews at 8:42 am, Oct 21, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>107985</u>	PRINTER SN <u>099.3586.820</u>	DATE OF INSPECTION <u>10-15-22</u>
------------------------------------	-----------------------------------	---------------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) <u>200 North Ave. Sparta, MO 65753 (Sparta P.D. Booking Room)</u>	TIME OF INSPECTION <u>1226</u>
---	-----------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER GUTH LAB LOT # 21380 EXP. DATE 9/13/23

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.9°C SIMULATOR SN MP 3584 SIMULATOR EXP DATE 03/03/2023

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • <u>.101</u>	TEST 2 • <u>.103</u>	TEST 3 • <u>.102</u>
----------------------	----------------------	----------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(.0-.04) <u>1</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	(OVER .19) <u>0</u>
-------------------	-------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
NONE

INSPECTING OFFICER SIGNATURE 	PRINT NAME <u>LOREN NYSTROM</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>210032 2-23-23</u>	TELEPHONE NUMBER <u>417-242-5511</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
2875 James Boulevard
Poplar Bluff, MO 63901

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01328

Temp Date Time 210L
s/

Air Blank:
10/15/22 12:27 .000
Calibration Check:
22 10/15/22 12:27 .101

Subject Name

Test

Subject I.D.

#1

Operator Name, I.D.

L. Nystrom 615

Location

Sparta P.D. Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01329

Temp Date Time 210L
s/

Air Blank:
10/15/22 12:32 .000
Calibration Check:
22 10/15/22 12:32 .103

Subject Name

Test

Subject I.D.

#2

Operator Name, I.D.

L. Nystrom 615

Location

Sparta P.D. Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01331

Temp Date Time 210L
s/

Air Blank:
10/15/22 12:35 .000
Calibration Check:
23 10/15/22 12:35 .102

Subject Name

Test

Subject I.D.

#3

Operator Name, I.D.

L. Nystrom 615

Location

Sparta P.D. Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01332

Temp Date Time 210L
s/

VOID: RFI
12 10/15/22 12:36

Subject Name

Test

Subject I.D.

RFI

Operator Name, I.D.

L. Nystrom 615

Location

Sparta P.D. Booking

AS IV Serial no: 107985
Version no: 532B

TEST RECORD 01333

Temp Date Time 210L
s/

Air Blank:
10/15/22 12:39 .000
Calibration Check:
23 10/15/22 12:39 .000

Subject Name

Test Blank

Subject I.D.

Blank

Operator Name, I.D.

L. Nystrom 615

Location

Sparta P.D. Booking



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
LOREN NYSTROM

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/23/2021



DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210032

EXPIRES 2/23/2023



DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator NYSTROM, LOREN
Permit No 210032
Date Issued 2/23/2021 Date Expires 2/23/2023

