

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 21001

EXPIRATION DATE: June 16, 2023 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 21001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1222 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is June 17, 2021 The expiration date for this lot number is June 16, 2023 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.

AS IV Serial no: 107990
Version no: 532B

TEST RECORD 00841

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
09/30/22 15:38 .000
Calibration Check:
20 09/30/22 15:38 .099

Subject Name
MAINTENANCE 1
Subject I.D.

Operator Name, I.D.
K.E. KINDLE, 220045
Location
1503 E. US-24
MOBERLY, MO

AS IV Serial no: 107990
Version no: 532B

TEST RECORD 00842

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
09/30/22 15:41 .000
Calibration Check:
21 09/30/22 15:41 .100

Subject Name
MAINTENANCE 2
Subject I.D.

Operator Name, I.D.
K.E. KINDLE, 220045
Location
1503 E. US-24
MOBERLY, MO

AS IV Serial no: 107990
Version no: 532B

TEST RECORD 00843

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
09/30/22 15:45 .000
Calibration Check:
22 09/30/22 15:45 .100

Subject Name
MAINTENANCE 3
Subject I.D.

Operator Name, I.D.
K.E. KINDLE, 220045
Location
1503 E. US-24
MOBERLY, MO

AS IV Serial no: 107990
Version no: 532B

TEST RECORD 00844

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 09/30/22 15:49

Subject Name
RFI TEST
Subject I.D.

Operator Name, I.D.
K.E. KINDLE, 220045
Location
1503 E. US-24
MOBERLY, MO



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KEEGAN E. KINDLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/10/2022

NUMBER 220045

EXPIRES 2/10/2024

Laura Q. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Keegan E. Kindle

, acting director
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator KINDLE, KEEGAN
Permit No 220045
Date Issued 2/10/2022 **Date Expires** 2/10/2024

