

RECEIVED

By Tracy Crews at 3:07 pm, Apr 18, 2022

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 107996	NAME OF AGENCY GLADSTONE PD	DATE OF INSPECTION 04/14/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 7010 N HOLMES ST GLADSTONE, MO 64118		TIME OF INSPECTION 9:59 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER INTOMIMETERS LOT # AG102503 EXP. DATE 01/25/2023 SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .102

TEST 2 .102

TEST 3 .102

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS 0 | (0-.04) 0 | (.05-.09) 0 | (.10-.14) 0 | (.15-.19) 0 | (OVER .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

ALL MAINTENANCE CONDUCTED WITHIN DHSS STANDARDS.

INSPECTING OFFICER

SIGNATURE

PRINT NAME

CHRISTOPHER BENNETT

TYPE II PERMIT NUMBER/EXPIRATION DATE

210203 09/14/2023

TELEPHONE NUMBER

(816) 436-3550

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01384

Temp Date Time ^{g/} 210L

Air Blank:
04/14/22 09:59 .000
Calibration Check:
23 04/14/22 09:59 .102

Subject Name

TEST 1

Subject I.D.

1

Operator Name, I.D.

BENNETT 210203

Location

7010 N HOLMES ST

GLADSTONE, MO 64118

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01385

Temp Date Time ^{g/} 210L

Air Blank:
04/14/22 10:01 .000
Calibration Check:
24 04/14/22 10:01 .102

Subject Name

TEST

Subject I.D.

2

Operator Name, I.D.

BENNETT 210203

Location

7010 N HOLMES ST

GLADSTONE, MO 64118

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01386

Temp Date Time ^{g/} 210L

Air Blank:
04/14/22 10:02 .000
Calibration Check:
25 04/14/22 10:02 .102

Subject Name

TEST

Subject I.D.

3

Operator Name, I.D.

BENNETT 210203

Location

7010 N HOLMES ST

GLADSTONE, MO 64118

AS IV Serial no: 107996
Version no: 532B

TEST RECORD 01387

Temp Date Time ^{g/} 210L

VOID: RFI
12 04/14/22 10:04

Subject Name

TEST

Subject I.D.

RFI

Operator Name, I.D.

BENNETT 210203

Location

7010 N HOLMES ST

GLADSTONE, MO 64118



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 27-Jan-2021

Lot # AG102503 Model 108caccd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
25-Jan-2023	108	Ethanol Nitrogen	0.100 ± 2% BrAC (272 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2021.01.29 13:36:13 -06:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CHRISTOPHER BENNETT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/14/2021

NUMBER 210203

EXPIRES 9/14/2023

LS 5500 (2-20-21)

[Signature]
 DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature]
 DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BENNETT, CHRISTOPHER
 Permit No 210203
 Date Issued 9/14/2021 Date Expires 9/14/2023

