



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108263	NAME OF AGENCY Woodson Terrace Police	DATE OF INSPECTION 03-16-2022
LOCATION OF INSTRUMENT (STREET AND CITY) 4305 Woodson Road, Woodson Terrace MO 63134		TIME OF INSPECTION 2205

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth LOT # 21190 EXP. DATE 6-8-2023 (Bot # 135)
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN DR3942 SIM. NIST EXP DATE 8-23-2022

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 \rightarrow .095	TEST 2 \rightarrow .097	TEST 3 \rightarrow .098
----------------------------------	----------------------------------	----------------------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 1	(0-.04) 0	(.05-.09) 0	(.10-.14) 0	(.15-.19) 0	(OVER .19) 0
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Time Adjusted

INSPECTING OFFICER	
SIGNATURE <i>PO. M. Anton DSN 113</i>	PRINT NAME Matthew Anton DSN 113
TYPE II PERMIT NUMBER/EXPIRATION DATE 200184 5-28-2022	TELEPHONE NUMBER (636) 474-9941

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00629

Temp Date Time 210L ^{s/}

Air Blank:
03/16/22 22:14 .000
Calibration Check:
26 03/16/22 22:14 .098

Subject Name
Anton 113

Subject I.D.
same

Operator Name, I.D.
self test 3

Location
WTPD

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00628

Temp Date Time 210L ^{s/}

Air Blank:
03/16/22 22:12 .000
Calibration Check:
25 03/16/22 22:12 .097

Subject Name
Anton 113

Subject I.D.
same

Operator Name, I.D.
self test 2

Location
WTPD

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00627

Temp Date Time 210L ^{s/}

Air Blank:
03/16/22 22:11 .000
Calibration Check:
25 03/16/22 22:11 .095

Subject Name
Anton 113

Subject I.D.
same

Operator Name, I.D.
self test 1

Location
WTPD

RFI

AS IV Serial no: 108263
Version no: 532B

TEST RECORD 00631

Temp Date Time 210L ^{s/}

VOID: RFI
12 03/16/22 22:17

Subject Name
Anton 113

Subject I.D.
same

Operator Name, I.D.
self test RFI

Location
WTPD



GUTH LABORATORIES, INC.

500 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **June 9, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **June 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

**Ted L. Pauley, President
GUTH LABORATORIES, INC.**

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: DR3942 Manufacturer: Guth
 Model Number: 2100
 Agency: WOODSON TERRACE PD
 Agency Address: 4305 WOODSON RD, WOODSON TERRACE, MO 63134

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
 Uncertainty: 0.02
 Date of Certification: 11/6/2020 Date of Expiration: 11/6/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.02	.04

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 8/23/2021
 Certification Expiration: 8/23/2022
 Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
 Certification No: DR3942_8232021



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
MATTHEW D ANTON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/28/2020

NUMBER 200184

EXPIRES 5/28/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ANTON, MATTHEW
Permit No 200184
Date Issued 5/28/2020 **Date Expires** 5/28/2022