

AS IV Serial no: 188267
Version no: 532B

TEST RECORD 00175

Temp Date Time 210L ^{g/}
Air Blank: 06/08/22 22:52 .000
Calibration Check: 24 06/08/22 22:52 .101

Subject Name
TEST #1

Subject I.D.

Operator Name, I.D.

CPL M.J. HANZAMAN

Location # 210065

TROOP A

AS IV Serial no: 188267
Version no: 532B

TEST RECORD 00176

Temp Date Time 210L ^{g/}
Air Blank: 06/08/22 22:54 .000
Calibration Check: 24 06/08/22 22:54 .102

Subject Name
TEST #2

Subject I.D.

Operator Name, I.D.

CPL M.J. HANZAMAN

Location # 210065

TROOP A

AS IV Serial no: 188267
Version no: 532B

TEST RECORD 00177

Temp Date Time 210L ^{g/}
Air Blank: 06/08/22 22:55 .000
Calibration Check: 25 06/08/22 22:55 .101

Subject Name
TEST #3

Subject I.D.

Operator Name, I.D.

CPL M.J. HANZAMAN

Location # 210065

TROOP A

AS IV Serial no: 188267
Version no: 532B

TEST RECORD 00178

Temp Date Time 210L ^{g/}
VOID: RFI
12 06/08/22 22:58

Subject Name
RFI TEST

Subject I.D.

Operator Name, I.D.

CPL M.J. HANZAMAN

Location # 210065

TROOP A



RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.

LOT NUMBER: 20001

EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:

RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.

Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

MATTHEW HANRAHAN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/6/2021

NUMBER 210063

EXPIRES 4/6/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HANRAHAN, MATTHEW
Permit No 210063
Date Issued 4/6/2021 Date Expires 4/6/2023