



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

**RECEIVED**  
 By Tracy Crews at 4:17 pm, Nov 02, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108270	NAME OF AGENCY Holden Police Department	DATE OF INSPECTION 10/31/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 110 W. 3rd Street, Holden		TIME OF INSPECTION 4:00 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 22080 EXP. DATE 03/07/2024
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN MP 3572 SIM. NIST EXP DATE 11/10/2022

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1  .103	TEST 2  .103	TEST 3  .103
--------------	--------------	--------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Time adjusted. Printer ribbon replaced.  
 Instrument working correctly within Dept. of Health Standards.

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT NAME Karl Van Vickle
TYPE II PERMIT NUMBER/EXPIRATION DATE 220152 / 05-25-2024	TELEPHONE NUMBER (816) 850-4154

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 108270  
Version no: 532B

TEST RECORD 00788

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
10/31/22 16:17 .000

Calibration Check:  
21 10/31/22 16:17 .000

Monthly Maint  
Subject Name

Blank Standard  
Subject I.D.

K. VanVickle 220152  
Operator Name, I.D.

110 W. 3<sup>rd</sup> St. Holden  
Location

MO 64040

Holden PD

AS IV Serial no: 108270  
Version no: 532B

TEST RECORD 00789

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 10/31/22 16:18

Monthly Maint RFI ✓  
Subject Name

RFI CHECK  
Subject I.D.

K. VanVickle 220152  
Operator Name, I.D.

110 W. 3<sup>rd</sup> St. Holden  
Location

MO 64040

Holden PD

AS IV Serial no: 108270  
Version no: 532B

TEST RECORD 00790

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
10/31/22 16:19 .000

Calibration Check:  
21 10/31/22 16:19 .103

Monthly Maint  
Subject Name

Test # 1  
Subject I.D.

K. VanVickle 220152  
Operator Name, I.D.

110 W. 3<sup>rd</sup> St. Holden  
Location

MO 64040

Holden PD

AS IV Serial no: 108270  
Version no: 532B

TEST RECORD 00791

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
10/31/22 16:21 .000

Calibration Check:  
21 10/31/22 16:21 .103

Monthly Maint  
Subject Name

Test # 2  
Subject I.D.

K. VanVickle 220152  
Operator Name, I.D.

110 W. 3<sup>rd</sup> St. Holden  
Location

MO 64040

Holden PD

AS IV Serial no: 108270  
Version no: 532B

TEST RECORD 00792

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
10/31/22 16:23 .000

Calibration Check:  
21 10/31/22 16:23 .103

Monthly Maint  
Subject Name

Test # 3  
Subject I.D.

K. VanVickle 220152  
Operator Name, I.D.

110 W. 3<sup>rd</sup> St.  
Location

Holden, MO 64040

Holden PD



## GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

### CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **22080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 9, 2022**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 7, 2024** at 11:59 PM.

When used in a calibrated Simulator, operating at  $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$ , this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L  $\pm$  3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President  
GUTH LABORATORIES, INC.

*NIST Traceability:*

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.*

*All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**KARL E. VANVICKLE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/25/2022

NUMBER 220152

EXPIRES 5/25/2024

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** VANVICKLE, KARL  
**Permit No** 220152  
**Date Issued** 5/25/2022 **Date Expires** 5/25/2024

