



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---|--|----------------------------------|
| ALCO SENSOR IV SN 108390 | NAME OF AGENCY LAWRENCE COUNTY SHERIFF'S OFFICE | DATE OF INSPECTION 01/13/2022 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 240 N. MAIN ST - MT. VERNON, MO 65712 | | TIME OF INSPECTION 12:49 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

| |
|--|
| <input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) |
| <input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C) |
| <input checked="" type="checkbox"/> PRINTER WORKING PROPERLY |
| <input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY |

BREATH ALCOHOL ACCURACY STANDARDS

| | |
|---|--|
| <input type="checkbox"/> SIMULATOR SOLUTION | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u> | LOT # <u>AG027903</u> EXP. DATE <u>09/30/2022</u> |
| <input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) | SIM. SN _____ SIM. NIST EXP DATE _____ |

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

| |
|---|
| <input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE |
| <input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE |
| <input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE |

| | | |
|---------------|---------------|---------------|
| TEST 1 ➔ .096 | TEST 2 ➔ .096 | TEST 3 ➔ .096 |
|---------------|---------------|---------------|

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | |
|----------|---------|-----------|-----------|-----------|------------|
| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) | (.15-.19) | (OVER .19) |
|----------|---------|-----------|-----------|-----------|------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

- Printer replaced by the Missouri Highway Safety Center. New printer serial # 210239923

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE | PRINT NAME RYAN DEVOST |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 210079 - 04/20/2023 | TELEPHONE NUMBER (417) 466-2131 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 108390
Version no: 532B

TEST RECORD 00460

Temp Date Time ^{g/} 210L

Air Blank:
01/13/22 12:49 .000
Calibration Check:
21 01/13/22 12:49 .096

Subject Name

Subject I.D.

Operator Name, I.D.

Location

TEST #
Ngz Dent

AS IV Serial no: 108390
Version no: 532B

TEST RECORD 00461

Temp Date Time ^{g/} 210L

Air Blank:
01/13/22 12:51 .000
Calibration Check:
22 01/13/22 12:51 .096

Subject Name

Subject I.D.

Operator Name, I.D.

Location

TEST # 2
Ngz Dent

AS IV Serial no: 108390
Version no: 532B

TEST RECORD 00462

Temp Date Time ^{g/} 210L

Air Blank:
01/13/22 12:53 .000
Calibration Check:
23 01/13/22 12:53 .096

Subject Name

Subject I.D.

Operator Name, I.D.

Location

TEST # 3
Ngz Dent

AS IV Serial no: 108390
Version no: 532B

TEST RECORD 00463

Temp Date Time ^{g/} 210L

VOID: RFI
12 01/13/22 12:54

Subject Name

Subject I.D.

Operator Name, I.D.

Location

RFI TEST
Ngz Dent

AS IV Serial no: 108390
Version no: 532B

TEST RECORD 00464

Temp Date Time ^{g/} 210L

Air Blank:
01/13/22 12:56 .000
Calibration Check:
24 01/13/22 12:56 .000

Subject Name

Subject I.D.

Operator Name, I.D.

Location

SOBER Sample
Ngz Dent



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 6-Oct-2020

Lot # AG027903 Model 55cacd

Exp. Date

5-Oct-2022

Cyl. Type

55

Component

Ethanol

Nitrogen

Certified Concentration

0.100 ± 2% BrAC (272 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.

EB0010581

Concentration

392.1 ppm

EB0010570

259.8 ppm

EB0010285

208.0 ppm

EB0010561

103.6 ppm

EB0010681

52.12 ppm

RGM Serial No.

EB0010603

Concentration

393.0 ppm

EB0010559

258.2 ppm

EB0010595

208.3 ppm

EB0010562

104.2 ppm

EB0010579

52.81 ppm

CRM Serial No.

CC434668

Concentration

800.0 ppm

CC234503

253.0 ppm

CRM Serial No.

0056649

Concentration

390.1 ppm

0056662

150.2 ppm

Analytical Method:

NDIR

Digitally signed by Quality Control
Date: 2020.10.07 15:09:30 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

RYAN DEVOST

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air, Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/20/2021

NUMBER 210079

EXPIRES 4/20/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (9-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **DEVOST, RYAN**
 Permit No **210079**
 Date Issued **4/20/2021** Date Expires **4/20/2023**