



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---|----------------------------|----------------------------------|
| ALCO SENSOR IV SN 108392 | PRINTER SN 099.3586.814 | DATE OF INSPECTION 06/22/2022 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 5490 5th St. Cottleville MO 63304 | | TIME OF INSPECTION 8:45 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 21080 EXP. DATE 03/08/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34C SIMULATOR SN MP4952 SIMULATOR EXP DATE 07/12/2022

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|---------------|---------------|---------------|
| TEST 1 ➡ .099 | TEST 2 ➡ .100 | TEST 3 ➡ .098 |
|---------------|---------------|---------------|

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | |
|----------|---|---------|-----------|-----------|---|-----------|------------|
| REFUSALS | 5 | (0-.04) | (.05-.09) | (.10-.14) | 1 | (.15-.19) | (OVER .19) |
|----------|---|---------|-----------|-----------|---|-----------|------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

| | |
|--|------------------------------------|
| INSPECTING OFFICER | |
| SIGNATURE <i>[Signature]</i> | PRINT NAME Richard Phillips |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 220110 04-14-2024 | TELEPHONE NUMBER (636) 498-6464 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Poplar Bluff, MO 63901

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00789
Temp Date Time ^{s/} 210L

Air Blank:
06/22/22 20:45 .000
Calibration Check:
22 06/22/22 20:45 .099

Subject Name
TEST 1
Subject I.D.
220110
Operator Name, I.D.
R PHILLIPS 408
Location
BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00790
Temp Date Time ^{s/} 210L

Air Blank:
06/22/22 20:47 .000
Calibration Check:
23 06/22/22 20:47 .100

Subject Name
TEST 2
Subject I.D.
220110
Operator Name, I.D.
R PHILLIPS 408
Location
BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00791
Temp Date Time ^{s/} 210L

Air Blank:
06/22/22 20:49 .000
Calibration Check:
24 06/22/22 20:49 .098

Subject Name
TEST 3
Subject I.D.
220110
Operator Name, I.D.
R PHILLIPS 408
Location
BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00792
Temp Date Time ^{s/} 210L

VOID: RFI
12 06/22/22 20:51

Subject Name
RFI TEST
Subject I.D.
220110
Operator Name, I.D.
R PHILLIPS 408
Location
BOOKING ROOM

AS IV Serial no: 108392
Version no: 532B

TEST RECORD 00793
Temp Date Time ^{s/} 210L

Air Blank:
06/22/22 20:52 .000
Subject Test: Auto
25 06/22/22 20:52 .000

Subject Name
ZERO TEST
Subject I.D.
220110
Operator Name, I.D.
R PHILLIPS 408
Location
BOOKING ROOM

FAX Cover Sheet

**To: Missouri Public Health & Senior Services Department
Attn: Brian Lutmer
Fax: 573-840-9139**

**From: Ofc. Rick Phillips, Type II, Permit #220110, expires
04/14/2024**

**Cottleville Police Department
5490 5th Street, Cottleville, MO 63304
Telephone: 636-498-6464
Fax: 636-498-6573**

Date Sent: 06-29-2022

Subject: ASIV Monthly Maintenance for Jun '22 – July '22

Remarks:

I can be reached at 636-498-6464 if any questions.

Thank you

Ofc. Rick Phillips DSN 468

richard.phillips@cityofcottleville.com


Missouri Department of Health and Senior Services

 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2986 VOICE 1-800-735-2468

Donald G. Kauerauf
 Director

Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP4952 **Manufacturer:** Guth
Model Number: 12V500
Agency: COTTLEVILLE PD
Agency Address: 5490 FIFTH STREET, COTTLEVILLE, MO 63304

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 11/6/2020 **Date of Expiration:** 11/6/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

| <u>Simulator Average</u> | <u>NIST Average</u> | <u>Combined Uncertainty</u> |
|--------------------------|---------------------|-----------------------------|
| 34.01 | 34.01 | .02 |

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 7/12/2021
Certification Expiration: 7/12/2022
Simulator testing technician: D. DEBOARD

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: MP4952_7122021

X *Brian Lutmer*

DHSS BAP Scientist Approving

**GUTH LABORATORIES, INC.**

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21080** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 10, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
RICHARD PHILLIPS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/14/2022

NUMBER 220110

EXPIRES 4/14/2024

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LA9-4 (R6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PHILLIPS, RICHARD
Permit No 220110
Date Issued 4/14/2022 Date Expires 4/14/2024

