



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|---|--|----------------------------------|
| ALCO SENSOR IV SN 111324 | NAME OF AGENCY Claycomo Police Department | DATE OF INSPECTION 07/25/2022 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 115 E. 69 Highway, Claycomo, MO 64119 | | TIME OF INSPECTION 2:30 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

| | |
|---|--|
| <input type="checkbox"/> SIMULATOR SOLUTION | <input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE |
| <input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Intoximeters</u> | LOT # <u>AG109003</u> EXP. DATE <u>03/12/2023</u> |
| <input type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) | SIM. SN _____ SIM. NIST EXP DATE _____ |

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

| | | |
|---------------|---------------|---------------|
| TEST 1 ← .097 | TEST 2 ← .098 | TEST 3 ← .098 |
|---------------|---------------|---------------|

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | | | | | | | |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|


List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Loaner instrument from UCMO Safety Center - Department unit out for repairs

Instrument operating within MO DOHSS standards and guidelines

RECEIVED
By Tracy Cross at 12:52 pm, Jul 26, 2022

INSPECTING OFFICER

| | |
|--|-------------------------------------|
| SIGNATURE  | PRINT NAME P.O. Jason A. Lederer |
| TYPE II PERMIT NUMBER/EXPIRATION DATE 210193 / 09/02/2023 | TELEPHONE NUMBER (816) 452-4613 |

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111324
Version no: 532B

TEST RECORD 01220

| Temp | Date | Time | 9/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
07/25/22 14:46 .000
Calibration Check:
22 07/25/22 14:46 .097

Subject Name

TEST # 1

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111324
Version no: 532B

TEST RECORD 01221

| Temp | Date | Time | 9/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
07/25/22 14:48 .000
Calibration Check:
22 07/25/22 14:48 .098

Subject Name

TEST # 2

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111324
Version no: 532B

TEST RECORD 01222

| Temp | Date | Time | 9/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
07/25/22 14:50 .000
Calibration Check:
23 07/25/22 14:50 .098

Subject Name

TEST # 3

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111324
Version no: 532B

TEST RECORD 01223

| Temp | Date | Time | 9/ 210L |
|------|------|------|------------|
|------|------|------|------------|

VOID: RFI
12 07/25/22 14:51

Subject Name

RFI Ck

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111324
Version no: 532B

TEST RECORD 01224

| Temp | Date | Time | 9/ 210L |
|------|------|------|------------|
|------|------|------|------------|

Air Blank:
07/25/22 14:53 .000
Subject Test: Auto
23 07/25/22 14:53 .000

Subject Name

SAMPLE TEST

Subject I.D.


Operator Name, I.D.

LEDENCE 118/162

Location

PERMIT # 210193

Exp DATE 09/02/2023



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 31-Mar-2021

Lot # AG109003 Model 108cacc

| | | | |
|---------------------------------|-------------------------|---|--|
| <u>Exp. Date</u> 31-Mar-2023 | <u>Cyl. Type</u> 108 | <u>Component</u> Ethanol Nitrogen | <u>Certified Concentration</u> 0.100 ± 2% BrAC (272 ppm) Balance |
|---------------------------------|-------------------------|---|--|

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

| <u>RGM Serial No.</u> | <u>Concentration</u> | <u>RGM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| EB0010581 | 392.1 ppm | EB0010603 | 393.0 ppm |
| EB0010570 | 259.8 ppm | EB0010559 | 258.2 ppm |
| EB0010285 | 208.0 ppm | EB0010595 | 208.3 ppm |
| EB0010561 | 103.6 ppm | EB0010562 | 104.2 ppm |
| EB0010681 | 52.12 ppm | EB0010579 | 52.81 ppm |

| <u>CRM Serial No.</u> | <u>Concentration</u> | <u>CRM Serial No.</u> | <u>Concentration</u> |
|-----------------------|----------------------|-----------------------|----------------------|
| CC727481 | 800.0 ppm | CC727493 | 390.0 ppm |
| CC727496 | 253.0 ppm | CC727498 | 150.0 ppm |

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.03.31 18:04:07 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

JASON A. LEDERER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/2/2021

NUMBER 210193

EXPIRES 9/2/2023

Laura E. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kamm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LEDERER, JASON

Permit No 210193

Date Issued 9/2/2021 Date Expires 9/2/2023

