

**RECEIVED**  
 P. 002  
 By Brian Lutmer at 4:20 pm, Jul 18, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111326	PRINTER SN 099.3586.615	DATE OF INSPECTION 07/13/2022
LOCATION OF INSTRUMENT (STREET AND CITY) Richland PD 201 South Chestnut Street Richland Missouri 65556		TIME OF INSPECTION 2:09 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG109702 EXP. DATE 04/07/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIMULATOR SN \_\_\_\_\_ SIMULATOR EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <input checked="" type="checkbox"/> 0.100	TEST 2 <input checked="" type="checkbox"/> 0.099	TEST 3 <input checked="" type="checkbox"/> 0.099
--	--	--

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument is operating within established limits.

<b>INSPECTING OFFICER</b>	
SIGNATURE <i>David L. Moser</i>	PRINT NAME David L. Moser
TYPE II PERMIT NUMBER/EXPIRATION DATE 200277 11/05/2022	TELEPHONE NUMBER (573) 480-4144

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
 2875 James Boulevard  
 Poplar Bluff, MO 63901

AS IV Serial no: 111326  
Version no: 532B

TEST RECORD 00999

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
07/13/22 14:09 .000  
Calibration Check:  
22 07/13/22 14:09 .000

Subject Name  
BLANK TEST  
Subject I.D.

Operator Name, I.D.

DMOSER 200277  
Location

RICHLAND PD

AS IV Serial no: 111326  
Version no: 532B

TEST RECORD 00998

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
07/13/22 14:10 .000  
Subject Test: Man  
22 07/13/22 14:10 .100

Subject Name  
TEST 1  
Subject I.D.

Operator Name, I.D.

DMOSER 200277  
Location

RICHLAND PD

AS IV Serial no: 111326  
Version no: 532B

TEST RECORD 00991

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
07/13/22 14:12 .000  
Subject Test: Man  
23 07/13/22 14:12 .099

Subject Name  
TEST 2  
Subject I.D.

Operator Name, I.D.

DMOSER 200277  
Location

RICHLAND PD

AS IV Serial no: 111326  
Version no: 532B

TEST RECORD 00982

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
07/13/22 14:14 .000  
Subject Test: Man  
23 07/13/22 14:14 .099

Subject Name  
TEST 3  
Subject I.D.

Operator Name, I.D.

DMOSER 200277  
Location

RICHLAND PD

AS IV Serial no: 111326  
Version no: 532B

TEST RECORD 00903

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 07/13/22 14:15

Subject Name  
RFI TEST  
Subject I.D.

Operator Name, I.D.


DMOSER 200377  
Location

RICHLAND PD





STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**DAVID L. MOSER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/5/2020

NUMBER 200277

EXPIRES 11/5/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MISSOURI 11/5/20

LAB 11/5/20

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri*

Operator MOSER, DAVID  
Permit No 200277  
Date Issued 11/5/2020 Date Expires 11/5/2022