

RECEIVED
 P. 002
 By Tracy Crews at 12:59 pm, Nov 10, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111326	NAME OF AGENCY Richland Police Department	DATE OF INSPECTION 11/09/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 201 S. Chestnut Ave Richland Mo. 65556		TIME OF INSPECTION 3:31 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG109702 EXP. DATE 04/07/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .098

TEST 2 → .097

TEST 3 → .097

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
 Unit is operating within established limits.

INSPECTING OFFICER

SIGNATURE

PRINT NAME
 David L. Moser

TYPE II PERMIT NUMBER/EXPIRATION DATE
 220256 11/09/2024

TELEPHONE NUMBER
 (573) 765-4144

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

1

→

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00940

Temp Date Time ^{u/} 210L

Air Blank:
11/09/22 15:31 .000
Subject Test: Auto
23 11/09/22 15:31 .000

Subject Name

BLANK TEST

Subject I.D.

Operator Name: I.D.

DMOSER 220256

Location

RICHLAND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00943

Temp Date Time ^{u/} 210L

Air Blank:
11/09/22 15:36 .000
Subject Test: Man
24 11/09/22 15:36 .000

Subject Name

TEST 1

Subject I.D.

Operator Name: I.D.

DMOSER 220256

Location

RICHLAND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00944

Temp Date Time ^{u/} 210L

Air Blank:
11/09/22 15:37 .000
Subject Test: Man
24 11/09/22 15:37 .000

Subject Name

TEST 2

Subject I.D.

Operator Name: I.D.

DMOSER 220256

Location

RICHLAND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00945

Temp Date Time ^{u/} 210L

Air Blank:
11/09/22 15:39 .000
Subject Test: Man
25 11/09/22 15:39 .000

Subject Name

TEST 3

Subject I.D.

Operator Name: I.D.

DMOSER 220256

Location

RICHLAND PD

AS IV Serial no: 111326
Version no: 532B

TEST RECORD 00946

Temp Date Time ^{u/} 210L

VOID: RFI
12 11/09/22 15:41

Subject Name

RFI TEST

Subject I.D.

Operator Name: I.D.

DMOSER 220256

Location

RICHLAND PD



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DAVID L. MOSER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 11/9/2022

NUMBER 220256

EXPIRES 11/9/2024

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)


LAB-4 (R6-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MOSER, DAVID
Permit No 220256
Date Issued 11/9/2022 Date Expires 11/9/2024



THE CITY OF RICHLAND, MISSOURI
RICHLAND POLICE DEPARTMENT

201 South Chestnut Street
Post Office Box # 798
Richland, Mo 65556

Gregg Sellers
Chief of Police

PHONE: (573) 765-4144

Fax: (573) 765-3093

FACSIMILE TRANSMISSION INFORMATION

Enclosed is a confidential transmittal, which may be law enforcement sensitive. The entire transmission is intended only for the person (s) and/or organization (s) listed here-in. If this message is received in error, please forward it to the correct person/organization immediately or destroy the entire transmission and contact us at once.

Your access to this information and its distribution may be regulated by law and is punishable as such.

Thank you for your cooperation.

FROM: RICHLAND POLICE DEPARTMENT

Sending Party/Officer: SGT DAVID MOSER

Total Pages Sent: Cover + 4


SENT TO: 11/09/2022 16:55

Organization: MOSS

Attention: _____

Fax Number: 573 840 9139

COMMENTS: MONTHLY MAINT



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph. (314) 533-3100

Fax. (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

Test Date: 8-Apr-2021

Lot # AG109702 Model 108cacc

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
7-Apr-2023	108	Ethanol	0.100 ± 2% BrAC (272 ppm)
		Nitrogen	Balance


Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.04.08 18:24:39 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: 
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07