



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111630	NAME OF AGENCY Miller County Sheriff's Office	DATE OF INSPECTION 05/23/2022
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 1999 Hwy 52 Tuscumbia, Missouri 65082	TIME OF INSPECTION 11:00 pm
---	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 21190 EXP. DATE 06/08/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN MP 2941 SIM. NIST EXP DATE 12/08/2022

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .099	TEST 3 ← .099
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	5	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE 	PRINT NAME Corporal Patrick
TYPE II PERMIT NUMBER/EXPIRATION DATE 200233 08/24/2022	TELEPHONE NUMBER (573) 369-2341

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111630
Version no: 532B

TEST RECORD 00763

Temp Date Time ^{s/} 210L

Air Blank:
05/23/22 23:00 .000
Calibration Check:
21 05/23/22 23:00 .100

Subject Name

Test #1

Subject I.D.

Pattick 220136

Operator Name, I.D.

Miller County

Location

AS IV Serial no: 111630
Version no: 532B

TEST RECORD 00764

Temp Date Time ^{s/} 210L

Air Blank:
05/23/22 23:02 .000
Calibration Check:
22 05/23/22 23:02 .099

Subject Name

Test #2

Subject I.D.

Pattick 220136

Operator Name, I.D.

Miller County

Location

AS IV Serial no: 111630
Version no: 532B

TEST RECORD 00765

Temp Date Time ^{s/} 210L

Air Blank:
05/23/22 23:03 .000
Calibration Check:
22 05/23/22 23:03 .099

Subject Name

Test #3

Subject I.D.

Pattick 220136

Operator Name, I.D.

Miller County

Location

AS IV Serial no: 111630
Version no: 532B

TEST RECORD 00766

Temp Date Time ^{s/} 210L

Air Blank:
05/23/22 23:05 .000
Subject Test: Auto
23 05/23/22 23:05 .000

Subject Name

Blank test

Subject I.D.

Pattick 220136

Operator Name, I.D.

Miller County

Location

AS IV Serial no: 111630
Version no: 532B

TEST RECORD 00767

Temp Date Time ^{s/} 210L

VOID: RFI
12 05/23/22 23:07

Subject Name

RFI

Subject I.D.

Pattick 220136

Operator Name, I.D.

Miller County

Location



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **June 9, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **June 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
SCOTT E. PATRICK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

Laura G. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220136

Paula J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 5/11/2024

MO 580-0771 (5-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator PATRICK, SCOTT
Permit No 220136
Date Issued 5/11/2022 **Date Expires** 5/11/2024

