



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111637	NAME OF AGENCY Gasconade County Sheriff's Office	DATE OF INSPECTION 08/16/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 119 E. 1st Street Hermann, Mo, 65041		TIME OF INSPECTION 1:04 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

- SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG18903 EXP. DATE 07/08/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
  - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
  - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
  - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 <b>0.98</b>	TEST 2 <b>0.97</b>	TEST 3 <b>0.97</b>
--------------------	--------------------	--------------------

- RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE <i>Dustin Toelke</i>	PRINT NAME Dustin Toelke
TYPE II PERMIT NUMBER/EXPIRATION DATE 220172/06/24/2024	TELEPHONE NUMBER (573) 486-2424

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111637  
Version no: 532B

TEST RECORD 00193

Temp	Date	Time	g/210L
Air Blank:	08/16/22	01:54	.000
Calibration Check:	23 08/16/22	01:54	.098

Air Blank:  
08/16/22 01:54 .000  
Calibration Check:  
23 08/16/22 01:54 .098

Subject Name

TEST 1

Subject I.D.

TEST 1

Operator Name, I.D.

DUSTIN TOEIKER220172

Location

119 E. 1st St

Hermann, MO 65041

AS IV Serial no: 111637  
Version no: 532B

TEST RECORD 00194

Temp	Date	Time	g/210L
Air Blank:	08/16/22	01:57	.000
Calibration Check:	24 08/16/22	01:57	.097

Air Blank:  
08/16/22 01:57 .000  
Calibration Check:  
24 08/16/22 01:57 .097

Subject Name

TEST 2

Subject I.D.

TEST 2

Operator Name, I.D.

DUSTIN TOEIKER220172

Location

119 E. 1st Street

Hermann, MO 65041

AS IV Serial no: 111637  
Version no: 532B

TEST RECORD 00195

Temp	Date	Time	g/210L
Air Blank:	08/16/22	02:00	.000
Calibration Check:	24 08/16/22	02:00	.097

Air Blank:  
08/16/22 02:00 .000  
Calibration Check:  
24 08/16/22 02:00 .097

Subject Name

TEST 3

Subject I.D.

TEST 3

Operator Name, I.D.

DUSTIN TOEIKER220172

Location

119 E. 1st Street

Hermann, MO 65041

AS IV Serial no: 111637  
Version no: 532B

TEST RECORD 00196

Temp	Date	Time	g/210L
VOID: RFI	12 08/16/22	02:02	

VOID: RFI  
12 08/16/22 02:02

Subject Name

TEST RFI

Subject I.D.

TEST RFI

Operator Name, I.D.

DUSTIN TOEIKER220172

Location

119 E. 1st Street

Hermann, MO 65041



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

**Test Date:** 8-Jul-2021

**Lot #** AG118903 **Model** 108cacd

**Exp. Date**

8-Jul-2023

**Cyl. Type**

108

**Component**

Ethanol

Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)

Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

**Analytical Method:**

NDIR

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**DUSTIN TOELKE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

NUMBER 220172

EXPIRES 6/24/2024

*Mike Massman*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nielson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** TOELKE, DUSTIN  
**Permit No** 220172  
**Date Issued** 6/24/2022 **Date Expires** 6/24/2024

