



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111637	NAME OF AGENCY Gasconade County Sheriff's Office	DATE OF INSPECTION 10/25/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 119 E. 1st Street Hermann, Mo 65041		TIME OF INSPECTION 3:44 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Intoximeters LOT # AG118903 EXP. DATE 07/08/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .101	TEST 2 .101	TEST 3 .100
--------------	--------------	--------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Dustin Toelke
TYPE II PERMIT NUMBER/EXPIRATION DATE 220172 06/24/2024	TELEPHONE NUMBER (573) 486-2424

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111637
Version no: 532B

TEST RECORD 00202

Temp	Date	Time	s/210L
------	------	------	--------

Air Blank:
10/25/22 15:44 .000
Calibration Check:
17 10/25/22 15:44 .101

Subject Name

TPST 1

Subject I.D.

TPST 1

Operator Name, I.D.

DUSTIN TOPIKE 220172

Location

119 E. 1st street

Hermann, MO 65041

AS IV Serial no: 111637
Version no: 532B

TEST RECORD 00203

Temp	Date	Time	s/210L
------	------	------	--------

Air Blank:
10/25/22 15:46 .000
Calibration Check:
18 10/25/22 15:46 .101

Subject Name

TPST 2

Subject I.D.

TPST 2

Operator Name, I.D.

DUSTIN TOPIKE 220172

Location

119 E. 1st street

Hermann, MO 65041

AS IV Serial no: 111637
Version no: 532B

TEST RECORD 00204

Temp	Date	Time	s/210L
------	------	------	--------

Air Blank:
10/25/22 15:49 .000
Calibration Check:
19 10/25/22 15:49 .100

Subject Name

TPST 3

Subject I.D.

TPST 3

Operator Name, I.D.

DUSTIN TOPIKE 220172

Location

119 E. 1st street

Hermann, MO 65041

AS IV Serial no: 111637
Version no: 532B

TEST RECORD 00205

Temp	Date	Time	s/210L
------	------	------	--------

VOID: RFI
12 10/25/22 15:51

Subject Name

TPST RFI

Subject I.D.

TPST RFI

Operator Name, I.D.

DUSTIN TOPIKE 220172

Location

119 E. 1st street

Hermann, MO 65041



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
DUSTIN TOELKE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

NUMBER 220172

EXPIRES 6/24/2024

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator TOELKE, DUSTIN
Permit No 220172
Date Issued 6/24/2022 **Date Expires** 6/24/2024

