



AS IV Serial no: 111643  
Version no: 532B  
TEST RECORD 00444 s/  
Temp Date Time 210L  
VOID: RPT  
12 04/06/72 23:54  
Subject Name  
*Monthly Maint*  
Subject I.D.  
Operator Name, I.D.  
*Dep. S. Holt*  
Location  
*GA*

AS IV Serial no: 111643  
Version no: 532B  
TEST RECORD 00443 s/  
Temp Date Time 210L  
Air Blank:  
04/06/72 23:52 .000  
Calibration Check:  
24 04/06/72 23:52 .100  
Subject Name  
*Monthly Maint*  
Subject I.D.  
Operator Name, I.D.  
*Dep. S. Holt*  
Location  
*GA*

AS IV Serial no: 111643  
Version no: 532B  
TEST RECORD 00442 s/  
Temp Date Time 210L  
Air Blank:  
04/06/72 23:51 .000  
Calibration Check:  
24 04/06/72 23:51 .100  
Subject Name  
*Monthly Maint*  
Subject I.D.  
Operator Name, I.D.  
*Dep. S. Holt*  
Location  
*GA*

AS IV Serial no: 111643  
Version no: 532B  
TEST RECORD 00441 s/  
Temp Date Time 210L  
Air Blank:  
04/06/72 23:49 .000  
Calibration Check:  
23 04/06/72 23:49 .100  
Subject Name  
*Monthly Maint*  
Subject I.D.  
Operator Name, I.D.  
*Dep. S. Holt*  
Location  
*GA*



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*

Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 1-Jun-2021

**Lot # AG114701 Model 108cacc**

**Exp. Date**

27-May-2023

**Cyl. Type**

108

**Component**

Ethanol  
Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)  
Balance

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

**RGM Serial No.**

EB0010581

**Concentration**

392.1 ppm

EB0010570

259.8 ppm

EB0010285

208.0 ppm

EB0010561

103.6 ppm

EB0010681

52.12 ppm

**RGM Serial No.**

EB0010603

**Concentration**

393.0 ppm

EB0010559

258.2 ppm

EB0010595

208.3 ppm

EB0010562

104.2 ppm

EB0010579

52.81 ppm

**CRM Serial No.**

CC434668

**Concentration**

800.0 ppm

CC234503

253.0 ppm

**CRM Serial No.**

0056649

**Concentration**

390.1 ppm

0056662

150.2 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2021.06.03 17:37:33 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**

**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**SEAN R. STOFF**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/11/2020

NUMBER 200304

EXPIRES 12/11/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MD 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** STOFF, SEAN  
**Permit No** 200304  
**Date Issued** 12/11/2020    **Date Expires** 12/11/2022

