



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111654	NAME OF AGENCY Platte County Sheriff's Office	DATE OF INSPECTION 09/15/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 415 Third Street, Platte City, Missouri 64079		TIME OF INSPECTION 7:06 pm

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters, Inc. LOT # AG114002 EXP. DATE 05/20/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .095

TEST 2 .095

TEST 3 .095

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	(0-.04)    2	(.05-.09)	(.10-.14)    1	(.15-.19)	(OVER .19)    1
----------	--------------	-----------	----------------	-----------	-----------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Taken out of service on 8/10/22 failed calibration checks during monthly maintenance. Taken to Missouri Safety Center and re-calibrated on 8/31/22

**INSPECTING OFFICER**

SIGNATURE

PRINT NAME

William Beeler

TYPE II PERMIT NUMBER/EXPIRATION DATE

220122/05-11-2024

TELEPHONE NUMBER

(816) 858-3521

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111654  
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00158

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/15/22 19:06 .000  
Calibration Check:  
23 09/15/22 19:06 .095

Subject Name

Test 1

Subject I.D.

Operator Name, I.D.

W. Beeler 220122  
5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 111654  
Version no: 532B

TEST RECORD 00159

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/15/22 19:11 .000  
Calibration Check:  
23 09/15/22 19:11 .095

Subject Name

Test 2

Subject I.D.

Operator Name, I.D.

W. Beeler 220122  
5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 111654  
Version no: 532B

TEST RECORD 00160

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/15/22 19:12 .000  
Calibration Check:  
23 09/15/22 19:12 .095

Subject Name

Test 3

Subject I.D.

Operator Name, I.D.

W. Beeler 220122  
5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 111654  
Version no: 532B

TEST RECORD 00161

Temp Date Time <sup>s/</sup> 210L

Air Blank:  
09/15/22 19:14 .000  
Subject Test: Auto  
24 09/15/22 19:14 .000

Subject Name

Blank Test

Subject I.D.

Operator Name, I.D.

W. Beeler 220122  
5/11/2024

Location

415 Third St

Platte City MO

AS IV Serial no: 111654  
Version no: 532B

TEST RECORD 00162

Temp Date Time <sup>s/</sup> 210L

VOID: RFI  
12 09/15/22 19:16

Subject Name

RFI

Subject I.D.

Operator Name, I.D.

W. Beeler 220122  
5/11/2024

Location

415 Third St

Platte City MO



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

Exclusive Supplier

Intoximeters, Inc.

2081 Craig Road

St. Louis, Mo 63146

**Test Date:** 21-May-2021

**Lot # AG114002 Model 108cadd**

**Exp. Date**

20-May-2023

**Cyl. Type**

108

**Component**

Ethanol

Nitrogen

**Certified Concentration**

0.100 ± 2% BrAC (272 ppm)

Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

**RGM Serial No.**

EB0010581

EB0010570

EB0010285

EB0010561

EB0010681

**Concentration**

392.1 ppm

259.8 ppm

208.0 ppm

103.6 ppm

52.12 ppm

**RGM Serial No.**

EB0010603

EB0010559

EB0010595

EB0010562

EB0010579

**Concentration**

393.0 ppm

258.2 ppm

208.3 ppm

104.2 ppm

52.81 ppm

**CRM Serial No.**

CC434668

CC234503

**Concentration**

800.0 ppm

253.0 ppm

**CRM Serial No.**

0056649

0056662

**Concentration**

390.1 ppm

150.2 ppm

**Analytical Method:**

NDIR

Digitally signed by Quality Control  
Date: 2021.05.21 11:39:32 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06

ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**WILLIAM BEELER**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

NUMBER 220122

EXPIRES 5/11/2024

*Laura P. Wang*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator BEELER, WILLIAM  
 Permit No 220122  
 Date Issued 5/11/2022 Date Expires 5/11/2024

