



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111655	NAME OF AGENCY Claycomo Police Department	DATE OF INSPECTION 06/03/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 115 E. 69 Highway, Claycomo, MO 64119		TIME OF INSPECTION 7:00 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG109003 EXP. DATE 03/31/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .095

TEST 2 ← .097

TEST 3 ← .097

RFI DETECTOR OPERATING


**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
 (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	1	(0-.04)	0	(.05-.09)	0	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use RECEIVED By Tracy Crews at 7:52 am, Jun 06, 2022 side if necessary).

Instrument operating within MO DOHSS standards and guidelines

INSPECTING OFFICER

SIGNATURE 	PRINT NAME P.O. Jason A. Lederer
TYPE II PERMIT NUMBER/EXPIRATION DATE 210193 / 09/02/2023	TELEPHONE NUMBER (816) 452-4613

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111655
Version no: 532B

TEST RECORD 01040

Temp	Date	Time	g/ 210L
------	------	------	------------

Air Blank:
06/03/22 19:13 .000
Calibration Check:
23 06/03/22 19:13 .095

Subject Name

TEST #1

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111655
Version no: 532B

TEST RECORD 01041

Temp	Date	Time	g/ 210L
------	------	------	------------

Air Blank:
06/03/22 19:15 .000
Calibration Check:
24 06/03/22 19:15 .097

Subject Name

TEST #2

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111655
Version no: 532B

TEST RECORD 01042

Temp	Date	Time	g/ 210L
------	------	------	------------

Air Blank:
06/03/22 19:17 .000
Calibration Check:
25 06/03/22 19:17 .097

Subject Name

TEST #3

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111655
Version no: 532B

TEST RECORD 01043

Temp	Date	Time	g/ 210L
------	------	------	------------

VOID: RFI
12 06/03/22 19:20

Subject Name

RFI CHECK

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111655
Version no: 532B

TEST RECORD 01044

Temp	Date	Time	g/ 210L
------	------	------	------------

Air Blank:
06/03/22 19:22 .000
Subject Test: Auto
25 06/03/22 19:22 .000

Subject Name

SAMPLE TEST

Subject I.D.


Operator Name, I.D.

LEDERER #118/162

Location

PERMIT #210193

EXP DATE 09/02/2023



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 31-Mar-2021

Lot # AG109003 Model 108cacd

<u>Exp. Date</u> 31-Mar-2023	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.100 ± 2% BrAC (272 ppm) Balance
---------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.03.31 18:04:07 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

JASON A. LEDERER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/2/2021

NUMBER 210193

EXPIRES 9/2/2023

Laura E. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kamm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LEDERER, JASON
 Permit No 210193
 Date Issued 9/2/2021 Date Expires 9/2/2023

