



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111659	PRINTER SN 09B.3589.478	DATE OF INSPECTION 05/10/2022
-----------------------------	----------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 4001 NE Lakewood Way Ct., Lee's Summit, Mo 64064	TIME OF INSPECTION 8:58 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG114701 EXP. DATE 05/27/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIMULATOR SN _____ SIMULATOR EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .099	TEST 2 ➡ .098	TEST 3 ➡ .098
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	8	(0-.04)	0	(.05-.09)	2	(.10-.14)	1	(.15-.19)	1	(OVER .19)	2
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS standards and guidelines.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME DEPUTY D. LOVE #61/0430
---------------	---------------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 220006 01-06-2024	TELEPHONE NUMBER (816) 524-4302
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office
 2875 James Boulevard
 Ponlar Bluff MO 63901

AS IV Serial no: 111659
Version no: 532B

TEST RECORD 00737

Temp Date Time 210L ^{g/}

Air Blank: 05/10/22 20:58 .000
Calibration Check: 23 05/10/22 20:58 .099

Subject Name

Monthly Maint.

Subject I.D.

Operator Name, I.D.

Dq. Dustin Love

Location

SSSO-GHQ

AS IV Serial no: 111659
Version no: 532B

TEST RECORD 00738

Temp Date Time 210L ^{g/}

Air Blank: 05/10/22 20:59 .000
Calibration Check: 24 05/10/22 20:59 .098

Subject Name

Monthly Maint.

Subject I.D.

Operator Name, I.D.

Dq. Dustin Love

Location

SSSO-GHQ

AS IV Serial no: 111659
Version no: 532B

TEST RECORD 00739

Temp Date Time 210L ^{g/}

Air Blank: 05/10/22 21:01 .000
Calibration Check: 25 05/10/22 21:01 .098

Subject Name

Monthly Maint.

Subject I.D.

Operator Name, I.D.

Dq. Dustin Love

Location

SSSO-GHQ

AS IV Serial no: 111659
Version no: 532B

TEST RECORD 00740

Temp Date Time 210L ^{g/}

VOID: RFI
12 05/10/22 21:02

Subject Name

Monthly Maint.

Subject I.D.

Operator Name, I.D.

Dq. Dustin Love

Location

SSSO-GHQ



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 1-Jun-2021

Lot # AG114701 Model 108cadd

Exp. Date 27-May-2023	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm) Balance
---------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<table border="0"> <tr> <th><u>RGM Serial No.</u></th> <th><u>Concentration</u></th> </tr> <tr> <td>EB0010581</td> <td>392.1 ppm</td> </tr> <tr> <td>EB0010570</td> <td>259.8 ppm</td> </tr> <tr> <td>EB0010285</td> <td>208.0 ppm</td> </tr> <tr> <td>EB0010561</td> <td>103.6 ppm</td> </tr> <tr> <td>EB0010681</td> <td>52.12 ppm</td> </tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010581	392.1 ppm	EB0010570	259.8 ppm	EB0010285	208.0 ppm	EB0010561	103.6 ppm	EB0010681	52.12 ppm	<table border="0"> <tr> <th><u>RGM Serial No.</u></th> <th><u>Concentration</u></th> </tr> <tr> <td>EB0010603</td> <td>393.0 ppm</td> </tr> <tr> <td>EB0010559</td> <td>258.2 ppm</td> </tr> <tr> <td>EB0010595</td> <td>208.3 ppm</td> </tr> <tr> <td>EB0010562</td> <td>104.2 ppm</td> </tr> <tr> <td>EB0010579</td> <td>52.81 ppm</td> </tr> </table>	<u>RGM Serial No.</u>	<u>Concentration</u>	EB0010603	393.0 ppm	EB0010559	258.2 ppm	EB0010595	208.3 ppm	EB0010562	104.2 ppm	EB0010579	52.81 ppm
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010581	392.1 ppm																								
EB0010570	259.8 ppm																								
EB0010285	208.0 ppm																								
EB0010561	103.6 ppm																								
EB0010681	52.12 ppm																								
<u>RGM Serial No.</u>	<u>Concentration</u>																								
EB0010603	393.0 ppm																								
EB0010559	258.2 ppm																								
EB0010595	208.3 ppm																								
EB0010562	104.2 ppm																								
EB0010579	52.81 ppm																								
<table border="0"> <tr> <th><u>CRM Serial No.</u></th> <th><u>Concentration</u></th> </tr> <tr> <td>CC434668</td> <td>800.0 ppm</td> </tr> <tr> <td>CC234503</td> <td>253.0 ppm</td> </tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	CC434668	800.0 ppm	CC234503	253.0 ppm	<table border="0"> <tr> <th><u>CRM Serial No.</u></th> <th><u>Concentration</u></th> </tr> <tr> <td>0056649</td> <td>390.1 ppm</td> </tr> <tr> <td>0056662</td> <td>150.2 ppm</td> </tr> </table>	<u>CRM Serial No.</u>	<u>Concentration</u>	0056649	390.1 ppm	0056662	150.2 ppm												
<u>CRM Serial No.</u>	<u>Concentration</u>																								
CC434668	800.0 ppm																								
CC234503	253.0 ppm																								
<u>CRM Serial No.</u>	<u>Concentration</u>																								
0056649	390.1 ppm																								
0056662	150.2 ppm																								

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2021.06.03 17:37:33 -05:00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

 Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
DUSTIN B. LOVE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 1/6/2022

Laura G. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220006

Donald A. Krumm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 1/6/2024

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator LOVE, DUSTIN
Permit No 220006
Date Issued 1/6/2022 **Date Expires** 1/6/2024

