



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

RECEIVED  
 03/17/2022 08:11:28 am, Room 17.2

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

|  |  |                                  |
|--|--|----------------------------------|
| ALCO SENSOR IV SN<br>111670  | NAME OF AGENCY<br>Vernon County Sheriff's Office | DATE OF INSPECTION<br>03/17/2022 |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br>2040 E. Hunter St. Nevada, Missouri, 64772 |  | TIME OF INSPECTION<br>8:05 am    |

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

**BREATH ALCOHOL ACCURACY STANDARDS**

SIMULATOR SOLUTION  COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG019902 EXP. DATE 07/17/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) \_\_\_\_\_ SIM. SN \_\_\_\_\_ SIM. NIST EXP DATE \_\_\_\_\_

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)  
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

|              |              |              |
|--------------|--------------|--------------|
| TEST 1  .079 | TEST 2  .079 | TEST 3  .079 |
|--------------|--------------|--------------|

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

|          |   |         |   |           |   |           |   |           |   |            |   |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|
| REFUSALS | 0 | (0-.04) | 0 | (.05-.09) | 0 | (.10-.14) | 0 | (.15-.19) | 0 | (OVER .19) | 0 |
|----------|---|---------|---|-----------|---|-----------|---|-----------|---|------------|---|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Changed Paper

**INSPECTING OFFICER**

|  |                                    |
|--|------------------------------------|
| SIGNATURE<br>  | PRINT NAME<br>M. Weisensee         |
| TYPE II PERMIT NUMBER/EXPIRATION DATE<br>220096 03/16/2024 | TELEPHONE NUMBER<br>(417) 283-4400 |

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00320 s/

Temp Date Time 210L

VOID: RFI  
12 03/17/22 08:27

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00319 s/

Temp Date Time 210L

Air Blank:  
03/17/22 08:26 .000

Calibration Check:  
24 03/17/22 08:26 .079

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00318 s/

Temp Date Time 210L

Air Blank:  
03/17/22 08:25 .000

Calibration Check:  
24 03/17/22 08:25 .079

Subject Name

Subject I.D.

Operator Name, I.D.

Location

AS IV Serial no: 111670  
Version no: 532B

TEST RECORD 00317 s/

Temp Date Time 210L

Air Blank:  
03/17/22 08:23 .000

Calibration Check:  
23 03/17/22 08:23 .079

Subject Name

Subject I.D.

Operator Name, I.D.

Location



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**

*Exclusive Supplier*

Intoximeters, Inc.  
2081 Craig Road  
St. Louis, Mo 63146

**Test Date:** 21-Jul-2020

**Lot # AG019902 Model 108cacc**

|  |                                |  |  |
|--|--------------------------------|--|--|
| <b><u>Exp. Date</u></b><br>17-Jul-2022 | <b><u>Cyl. Type</u></b><br>108 | <b><u>Component</u></b><br>Ethanol<br>Nitrogen | <b><u>Certified Concentration</u></b><br>0.080 ± 0.002 BrAC (208 ppm)<br>Balance |
|--|--------------------------------|--|--|

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

|   |  |   |  |
|---|--|---|--|
| <b><u>RGM Serial No.</u></b><br>EB0010581<br>EB0010570<br>EB0010285<br>EB0010561<br>EB0010681 | <b><u>Concentration</u></b><br>392.1 ppm<br>259.8 ppm<br>208.0 ppm<br>103.6 ppm<br>52.12 ppm | <b><u>RGM Serial No.</u></b><br>EB0010603<br>EB0010559<br>EB0010595<br>EB0010562<br>EB0010579 | <b><u>Concentration</u></b><br>393.0 ppm<br>258.2 ppm<br>208.3 ppm<br>104.2 ppm<br>52.81 ppm |
| <b><u>CRM Serial No.</u></b><br>CC434668<br>CC234503  | <b><u>Concentration</u></b><br>800.0 ppm<br>253.0 ppm  | <b><u>CRM Serial No.</u></b><br>0056649<br>0056662  | <b><u>Concentration</u></b><br>390.1 ppm<br>150.2 ppm  |

**Analytical Method:** NDIR

Digitally signed by Quality Control  
Date: 2020.07.22 16:48:30 -05:00  
Reason: Dry gas standard certification of analysis  
Location: Airgas USA LLC (Lab)

Approved for Release:

Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT  
 TYPE II**

**MICHAEL WEISENSEE**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/16/2022

NUMBER 220096

EXPIRES 3/16/2024

*Laura Q. Way*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*David F. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** WEISENSEE, MICHAEL  
**Permit No** 220096  
**Date Issued** 3/16/2022 **Date Expires** 3/16/2024

