

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00228
Temp Date Time 2:01

Air Blank
05/25/72 16:45 .000
Calibration Check
23 05/25/72 16:45 .001

Subject Name
MD Green
940

Operator Name, I.D.
Mud Green 940

Location
Zone Office

Frequency
Monthly Maint

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00227
Temp Date Time 2:01

Air Blank
05/27/72 16:40 .000
Calibration Check
23 05/27/72 16:40 .000

Subject Name
MD Green
940

Operator Name, I.D.
Mud Green 940

Location
Zone Office

Frequency
Monthly Maint

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00226
Temp Date Time 2:01

Air Blank
05/25/72 16:47 .000
Calibration Check
23 05/25/72 16:47 .001

Subject Name
MD Green
940

Operator Name, I.D.
Mud Green 940

Location
Zone Office

Frequency
Monthly Maint

AS IV Serial no: 111672
Version no: 5328

TEST RECORD 00225
Temp Date Time 2:01

Air Blank
05/25/72 16:45 .000
Calibration Check
23 05/25/72 16:45 .001

Subject Name
MD Green
940

Operator Name, I.D.
Mud Green 940

Location
Zone Office

Frequency
Monthly Maint

RepCo Marketing Co
3101-188 Stony Brook Driv
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
MARK D. GREEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/9/2021

NUMBER 210199

EXPIRES 9/9/2023

Laura Q. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Donald A. Kamm

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator GREEN, MARK
 Permit No 210199
 Date Issued 9/9/2021 Date Expires 9/9/2023

