

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00407

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
08/06/22 10:34 .000
Calibration Check:
23 08/06/22 10:34 .096

Subject Name

Maintenance

Subject I.D.

Tpr. Ott

Operator Name, I.D.

Ray Co ZO
Location

#1

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00408

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
08/06/22 10:36 .000
Calibration Check:
23 08/06/22 10:36 .096

Subject Name

Maintenance

Subject I.D.

Tpr. Ott

Operator Name, I.D.

Ray Co ZO
Location

#2

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00409

Temp	Date	Time	s/ 210L
------	------	------	------------

Air Blank:
08/06/22 10:37 .000
Calibration Check:
24 08/06/22 10:37 .096

Subject Name

Maintenance

Subject I.D.

Tpr. Ott

Operator Name, I.D.

Ray Co ZO
Location

#3

AS IV Serial no: 111674
Version no: 532B

TEST RECORD 00406

Temp	Date	Time	s/ 210L
------	------	------	------------

VOID: RFI
12 08/06/22 10:32

Subject Name

Subject I.D.

Tpr. Ott

Operator Name, I.D.

Ray Co ZO
Location

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
NICHOLAS OTT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/16/2022
NUMBER 220088
EXPIRES 3/16/2024

[Signature]
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

[Signature: David F. Nicholson]
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MS-AS 07/1 (6.12)

LAB-015-12

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **OTT, NICHOLAS**
Permit No **220088**
Date Issued **3/16/2022** Date Expires **3/16/2024**

