

RECEIVED

By Tracy Crews at 8:24 am, Aug 09, 2022

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|-----------------------------------------------------------------------------------|------------------------------------------|----------------------------------|
| ALCO SENSOR IV SN 111735 | NAME OF AGENCY Linn Police Department | DATE OF INSPECTION 08/02/2022 |
| LOCATION OF INSTRUMENT (STREET AND CITY) 1200 East Main Street, Linn, MO 65051 | | TIME OF INSPECTION 2:21 pm |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER RepCo Marketing Co. LOT # 20001 EXP. DATE 10/07/2022 SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIM. SN MP6021 SIM. NIST EXP DATE 09/03/2022 CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ● .099

TEST 2 ● .100

TEST 3 ● .099

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS

(0-.04)

(.05-.09)

(.10-.14)

(.15-.19)

(OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Liquid changes. New liquid bottle #0426.

INSPECTING OFFICER

SIGNATURE

PRINT NAME

Michael A. Bickell

TYPE II PERMIT NUMBER/EXPIRATION DATE

TELEPHONE NUMBER

210204, 09/14/2023

(573) 897-2236

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111735
Version no: 532B

TEST RECORD 00217

Temp Date Time ^{a/} 210L

Air Blank:
08/02/22 14:21 .000
Calibration Check:
18 08/02/22 14:21 .099

Subject Name

Maint Test

Subject I.D.

#1

Operator Name, I.D.

M. A. Bickell 210204

Location

Linn Police Dept.

AS IV Serial no: 111735
Version no: 532B

TEST RECORD 00218

Temp Date Time ^{a/} 210L

Air Blank:
08/02/22 14:22 .000
Calibration Check:
19 08/02/22 14:22 .100

Subject Name

Maint Test

Subject I.D.

#2

Operator Name, I.D.

M. A. Bickell 210204

Location

Linn Police Dept.

AS IV Serial no: 111735
Version no: 532B

TEST RECORD 00219

Temp Date Time ^{a/} 210L

Air Blank:
08/02/22 14:24 .000
Calibration Check:
19 08/02/22 14:24 .099

Subject Name

Maint Test

Subject I.D.

#3

Operator Name, I.D.

M. A. Bickell 210204

Location

Linn Police Dept.

AS IV Serial no: 111735
Version no: 532B

TEST RECORD 00220

Temp Date Time ^{a/} 210L

VOID: RFI
12 08/02/22 14:29

Subject Name

Maint Test

Subject I.D.

RFI

Operator Name, I.D.

M. A. Bickell 210204

Location

Linn Police Dept.


STATE OF MISSOURI)
)
COUNTY OF OSAGE) SS

AFFIDAVIT

Before me, the undersigned authority, personally appeared Michael A. Bickell, who being by me duly sworn, deposed as follows:

My name is Michael A. Bickell, I am of sound mind, capable of making this affidavit, and personally acquainted with the facts herein stated:

I am the custodian of the records of the maintenance report for the portable Alco-Sensor IV, serial number 111735, with printer. Attached hereto are 2 pages of records beginning 08/02/2022. These 2 pages of records are kept by the Linn Police Department in the regular course of business, and it was the regular course of business of the Linn Police Department for an employee or representative of the Linn Police Department with knowledge of the act recorded to make the record or to transmit information thereof to be included in such record; and the record was made at or near the time of the act. The records attached hereto are the original or exact duplicates of the original



Michael A. Bickell, Affiant
Chief of Police, City of Linn Police Department

In witness whereof I have hereunto subscribed my name and affixed official my seal this 8th day of August, 2022.



Notary Public

My commission expires: March 20, 2024
Commissioned in Osage County, Missouri.

CARRIE L. GRELLNER
Notary Public - Notary Seal
State of Missouri
Commissioned for Osage County
My Commission Expires: March 20, 2024
Commission Number: 12317265