

RECEIVED

By Tracy Crews at 11:09 am, Jun 15, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111750	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 06/13/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 504 SE Blue Parkway, Lee's Summit, Missouri 64063		TIME OF INSPECTION 13:15

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)	
<input checked="" type="checkbox"/> TEMPERATURE OF ALCO SENSOR (10°C - 40°C)	
<input checked="" type="checkbox"/> PRINTER WORKING PROPERLY	
<input checked="" type="checkbox"/> TIME AND DATE DISPLAYING PROPERLY	

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>RepCo Marketing Co</u> LOT # <u>20001</u> EXP. DATE <u>10/07/2022</u>	
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34.01</u> SIM. SN <u>MP2206</u> SIM. NIST EXP DATE <u>12/29/2022</u>	
<input checked="" type="checkbox"/> CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT) Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)	
<input checked="" type="checkbox"/> 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE	
<input type="checkbox"/> 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE	
<input type="checkbox"/> 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE	

TEST 1 \leftarrow .102	TEST 2 \leftarrow .102	TEST 3 \leftarrow .103
--------------------------	--------------------------	--------------------------

<input checked="" type="checkbox"/> RFI DETECTOR OPERATING	
--	--

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 3	(0-.04) 0	(.05-.09) 1	(.10-.14) 1	(.15-.19) 3	(OVER .19) 1
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE <i>Cpl. A.A. Engelhart #247</i>	PRINT NAME Cpl. A.A. Engelhart #247
TYPE II PERMIT NUMBER/EXPIRATION DATE 200223 08/13/2022	TELEPHONE NUMBER (816) 622-0800

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111750
Version no: 532B

TEST RECORD 00495

Temp Date Time ^{s/} 210L

Air Blank:
06/13/22 13:18 .000
Calibration Check:
22 06/13/22 13:18 .102

Subject Name

Maint Test 1

Subject I.D.

Operator Name, I.D.

Cpl. A.A. Engellist 200223

Location

Troop A

AS IV Serial no: 111750
Version no: 532B

TEST RECORD 00496

Temp Date Time ^{s/} 210L

Air Blank:
06/13/22 13:20 .000
Calibration Check:
23 06/13/22 13:20 .102

Subject Name

Maint Test 2

Subject I.D.

Operator Name, I.D.

Cpl. A.A. Engellist 200223

Location

Troop A

AS IV Serial no: 111750
Version no: 532B

TEST RECORD 00497

Temp Date Time ^{s/} 210L

Air Blank:
06/13/22 13:22 .000
Calibration Check:
23 06/13/22 13:22 .103

Subject Name

Maint Test 3

Subject I.D.

Operator Name, I.D.

Cpl. A.A. Engellist 200223

Location

Troop A

AS IV Serial no: 111750
Version no: 532B

TEST RECORD 00498

Temp Date Time ^{s/} 210L

VOID: RFI
12 06/13/22 13:23

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

Cpl. A.A. Engellist 200223

Location

Troop A

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

**PERMIT
TYPE II**

AARON A ENGELHART

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/13/2020

NUMBER 200223

EXPIRES 8/13/2022

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (15-10)

MO 580 0771 (6-10)



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator ENGELHART, AARON
Permit No 200223
Date Issued 8/13/2020 Date Expires 8/13/2022