



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

RECEIVED
By: [illegible]
Date: 4/15/22

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>111752</u>	NAME OF AGENCY <u>STAFFORDS POLICE DEPT</u>	DATE OF INSPECTION <u>4-15-22</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>125 S. JEFFERSON STAFFORDS, MO</u>		TIME OF INSPECTION <u>1635</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

☐ DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

☒ TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

☒ PRINTER WORKING PROPERLY

☒ TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

☒ SIMULATOR SOLUTION ☐ COMPRESSED ETHANOL-GAS MIXTURE

☒ STANDARD SUPPLIER GUTH LABS LOT # 21380 EXP. DATE 9-13-23

☒ SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.0 SIM. SN 225379 SIM. NIST EXP DATE 5-27-22

☒ CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 • .100 TEST 2 • .100 TEST 3 • .100

☒ RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

AT TIME OF MAINTENANCE, INSTRUMENT WAS OPERATING IN ACCORDANCE WITH DHSS STANDARDS

INSPECTING OFFICER

SIGNATURE <u>[Signature]</u>	PRINT NAME <u>KRISTEN J. MALLORY</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>210135 EXP 6-30-23</u>	TELEPHONE NUMBER <u>417-736-4000</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00709

Temp Date Time 210L

Air Blank:
04/15/22 16:35 .000
Calibration Check:
21 04/15/22 16:35 .100

Subject Name

TEST #1

Subject I.D.

KRISTEN MALLORY

Operator Name, I.D.

210135 EXP 6-30-23

Location

125 S. JEFFERSON

STRAFFORD, MO

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00710

Temp Date Time 210L

Air Blank:
04/15/22 16:37 .000
Calibration Check:
22 04/15/22 16:37 .100

Subject Name

TEST #2

Subject I.D.

KRISTEN MALLORY

Operator Name, I.D.

210135 EXP 6-30-23

Location

125 S. JEFFERSON

STRAFFORD, MO

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00711

Temp Date Time 210L

Air Blank:
04/15/22 16:38 .000
Calibration Check:
22 04/15/22 16:38 .100

Subject Name

TEST #3

Subject I.D.

KRISTEN MALLORY

Operator Name, I.D.

210135 EXP 6-30-23

Location

125 S. JEFFERSON

STRAFFORD, MO

AS IV Serial no: 111752
Version no: 532B

TEST RECORD 00712

Temp Date Time 210L

VOID: RFI
12 04/15/22 16:39

Subject Name

RFI

Subject I.D.

KRISTEN MALLORY

Operator Name, I.D.

210135 EXP 6-30-23

Location

125 S. JEFFERSON

STRAFFORD, MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21380** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 15, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 13, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L +/- 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



**PERMIT
TYPE II**

KRISTEN J. MALLORY

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/30/2021

NUMBER 210135

EXPIRES 6/30/2023

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (5-10)

LAB-4 (R5-10)

