



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 9:45 am, May 10, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111754	NAME OF AGENCY MSHP	DATE OF INSPECTION 05/02/2022
-----------------------------	------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 102 W N MAIN ST. RICHMOND MO 64085	TIME OF INSPECTION 10:02 pm
--	--------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER REPCO MARKETING CO LOT # 20001 EXP. DATE 10/07/2022

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN MP2125 SIM. NIST EXP DATE 12/29/2022

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100

TEST 2 .100

TEST 3 .099

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE

PRINT NAME
 TPR. N. T. OTT

TYPE II PERMIT NUMBER/EXPIRATION DATE
 220088 03/16/2024

TELEPHONE NUMBER
 (816) 622-0800

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/-0.003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/-3% gms/210L breath when heated to 34 Degrees Celsius +/-0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00248

Temp Date Time ^{a/} 210L

Air Blank:
05/02/22 22:32 .000
Calibration Check:
22 05/02/22 22:32 .100

Subject Name
Maintenance

Subject I.D.

Tr. N. T. Ott

Operator Name, I.D.

220088 3/16/24

Location

Ray Co Zone Office

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00249

Temp Date Time ^{a/} 210L

Air Blank:
05/02/22 22:34 .000
Calibration Check:
22 05/02/22 22:34 .100

Subject Name

Maintenance

Subject I.D.

Tr. N. T. Ott

Operator Name, I.D.

220088 3/16/24

Location

Ray Co Zone Office

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00250

Temp Date Time ^{a/} 210L

Air Blank:
05/02/22 22:36 .000
Calibration Check:
22 05/02/22 22:36 .099

Subject Name

Maintenance

Subject I.D.

Tr. N. T. Ott

Operator Name, I.D.

220088 3/16/24

Location

Ray Co Zone Office

AS IV Serial no: 111754
Version no: 532B

TEST RECORD 00251

Temp Date Time ^{a/} 210L

NOISE PFI
12 05/02/22 22:37

Subject Name

Subject I.D.

Operator Name, I.D.

Location

Ray Co Zone Office

965



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
NICHOLAS OTT

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/16/2022

Laura P. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220088

David J. Nickelson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 3/16/2024

MO 580-0771 (6-10)

LAB-4 (R6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator OTT, NICHOLAS
Permit No 220088
Date Issued 3/16/2022 **Date Expires** 3/16/2024

