



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111759	NAME OF AGENCY CLAY COUNTY SHERIFF'S OFFICE	DATE OF INSPECTION 07/01/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 12 S. WATER STREET, LIBERTY, MO 64068		TIME OF INSPECTION 5:50 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS LOT # AG102503 EXP. DATE 01/25/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .099

TEST 2 ← .099

TEST 3 ← .099

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	1	(.10-.14)	1	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

REPLACE UNIT BATTERY

INSPECTING OFFICER

SIGNATURE

[Signature] #8199

PRINT NAME

JAMES BUSH

TYPE II PERMIT NUMBER/EXPIRATION DATE

210324 - 12/21/2023

TELEPHONE NUMBER

(816) 407-3702

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111759
Version no: 532B

TEST RECORD 00178

Temp Date Time ^{s/} 210L

Air Blank:
07/01/22 05:50 .000
Calibration Check:
21 07/01/22 05:50 .099

Subject Name

Monthly Maintenance

Subject I.D.

Test 1

Operator Name, I.D.

J. Bush #8199

Location

12 S. Water St.,

Liberty, MO 64068

AS IV Serial no: 111759
Version no: 532B

TEST RECORD 00179

Temp Date Time ^{s/} 210L

Air Blank:
07/01/22 05:52 .000
Calibration Check:
22 07/01/22 05:52 .099

Subject Name

Monthly Maintenance

Subject I.D.

Test 2

Operator Name, I.D.

J. Bush #8199

Location

12 S. Water St.,

Liberty, MO 64068

AS IV Serial no: 111759
Version no: 532B

TEST RECORD 00180

Temp Date Time ^{s/} 210L

Air Blank:
07/01/22 05:53 .000
Calibration Check:
22 07/01/22 05:53 .099

Subject Name

Monthly Maintenance

Subject I.D.

Test 3

Operator Name, I.D.

J. Bush #8199

Location

12 S. Water St.,

Liberty, MO 64068

AS IV Serial no: 111759
Version no: 532B

TEST RECORD 00181

Temp Date Time ^{s/} 210L

VOID: RFI
12 07/01/22 05:55

Subject Name

Monthly Maintenance

Subject I.D.

RFI Test

Operator Name, I.D.

J. Bush #8199

Location

12 S. Water St.,

Liberty, MO 64068



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
JAMES S. BUSH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2021

Laura E. Wray

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210324

Donald A. Ramsey

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 12/21/2023

10 580-0771 (6-10)

LAB-4, (16-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator BUSH, JAMES
Permit No 210324
Date issued 12/21/2021 Date Expires 12/21/2023

