

AS IV Serial no: 111762
Version no: 532B

TEST RECORD 00431

Temp Date Time 210L

Air Blank:
05/13/22 05:44 .000
Calibration Check:
21 05/13/22 05:44 .000

Subject Name

Monthly Maint
Subject I.D.

Self Test

Operator Name, I.D.

K. Stults 220094
Location

Stone Co Jail

AS IV Serial no: 111762
Version no: 532B

TEST RECORD 00432

Temp Date Time 210L

Air Blank:
05/13/22 05:48 .000
Calibration Check:
22 05/13/22 05:48 .102

Subject Name

Monthly Maint
Subject I.D.

Test 1

Operator Name, I.D.

K. Stults 220094
Location

Stone Co Jail

AS IV Serial no: 111762
Version no: 532B

TEST RECORD 00433

Temp Date Time 210L

Air Blank:
05/13/22 05:51 .000
Calibration Check:
23 05/13/22 05:51 .101

Subject Name

Monthly Maint
Subject I.D.

Test 2

Operator Name, I.D.

K. Stults 220094
Location

Stone Co Jail

AS IV Serial no: 111762
Version no: 532B

TEST RECORD 00434

Temp Date Time 210L

Air Blank:
05/13/22 05:54 .000
Calibration Check:
23 05/13/22 05:54 .101

Subject Name

Monthly Maint
Subject I.D.

Test 3

Operator Name, I.D.

K. Stults 220094
Location

Stone Co Jail

AS IV Serial no: 111762
Version no: 532B

TEST RECORD 00435

Temp Date Time 210L

VOID: RFI
12 05/13/22 05:56

Subject Name

Monthly Maint
Subject I.D.

RFI

Operator Name, I.D.

K. Stults 220094
Location

Stone Co Jail

Monthly Maint
05/13/2022
Stone County Jail
Dep Kyle Stults
Permit # 220094
Exp: 03/16/2024



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21190** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **June 9, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1206%** (w/vol) ethyl alcohol. The expiration date for this lot number is **June 8, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
KYLE G. STULTS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/16/2022

NUMBER 220094

EXPIRES 3/16/2024

Laura P. Way

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator STULTS, KYLE
 Permit No 220094
 Date Issued 3/16/2022 Date Expires 3/16/2024

