

AS IV Serial no: 111768
Version no: 532B

TEST RECORD - REPRINT

TEST RECORD 00663 s/
Temp Date Time 210L

Air Blank: 04/20/22 15:20 .000
Calibration Check: 20 04/20/22 15:20 .100

Subject Name
TEST 1
Subject I.D.

Operator Name, I.D.
D W HENLEY #727
Location

815 ORCHARD BLVD

FAIR GROVE, MO

WJ #727

AS IV Serial no: 111768
Version no: 532B

TEST RECORD 00664 s/
Temp Date Time 210L

Air Blank: 04/20/22 15:24 .000
Calibration Check: 21 04/20/22 15:24 .100

Subject Name
TEST 2
Subject I.D.

Operator Name, I.D.
D W HENLEY #727
Location

815 ORCHARD BLVD

FAIR GROVE, MO

WJ #727

AS IV Serial no: 111768
Version no: 532B

TEST RECORD 00665 s/
Temp Date Time 210L

Air Blank: 04/20/22 15:25 .000
Calibration Check: 22 04/20/22 15:25 .099

Subject Name
TEST 3
Subject I.D.

Operator Name, I.D.
D W HENLEY #727
Location

815 ORCHARD BLVD

FAIR GROVE, MO

WJ #727

AS IV Serial no: 111768
Version no: 532B

TEST RECORD 00666 s/

Temp Date Time 210L
VOID: RFI
12 04/20/22 15:26

Subject Name
RFI
Subject I.D.

Operator Name, I.D.
D W HENLEY #727
Location

815 ORCHARD BLVD

FAIR GROVE, MO

WJ #727



RepCo Marketing Co
3101-188 Stony Brook Drive
Raleigh, NC 27604
919-876-5480

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing Co.
LOT NUMBER: 20001
EXPIRATION DATE: October 7, 2022 at 11:59 p.m.

RepCo Marketing Co. certifies the following:


RepCo Marketing Co. prepared, tested and supplied Lot Number 20001 of Alcohol Certified Solution for simulators. This solution was manufactured and tested by RepCo Marketing Co., with confirmation by ISO 17025 and ISO 17034 accredited institution Alcohol Countermeasure Systems, using NIST standards. Random samples were analyzed by Alcohol Countermeasure Systems utilizing a gas chromatograph and found to contain .1227 gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is October 8, 2020 The expiration date for this lot number is October 7, 2022 at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.


Alma Palmer, Operations Manager
RepCo Marketing Co.

