



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111774	NAME OF AGENCY Carroll County Sheriff's Office	DATE OF INSPECTION 06/04/2022
-----------------------------	---	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 106 S. Folger St, Carrollton	TIME OF INSPECTION 2:22 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Labs LOT # 21380 EXP. DATE 09/13/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIM. SN MP3879 SIM. NIST EXP DATE 05/02/2023

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .101	TEST 3 ← .100
---------------	---------------	---------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE <i>Brian C. Woods</i>	PRINT NAME Brian C. Woods
------------------------------------	------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 200226 08/13/2022	TELEPHONE NUMBER (660) 542-2200
--	------------------------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **21380** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **September 15, 2021**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1214%** (w/vol) ethyl alcohol. The expiration date for this lot number is **September 13, 2023** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L** \pm 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00325

Temp	Date	Time	s/
		210L	

Air Blank:
06/04/22 14:22 .000
Calibration Check:
22 06/04/22 14:22 .100

Subject Name

Test #1

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00326

Temp	Date	Time	s/
		210L	

Air Blank:
06/04/22 14:24 .000
Calibration Check:
22 06/04/22 14:24 .101

Subject Name

Test #2

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00327

Temp	Date	Time	s/
		210L	

Air Blank:
06/04/22 14:27 .000
Calibration Check:
23 06/04/22 14:27 .100

Subject Name

Test #3

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00329

Temp	Date	Time	s/
		210L	

VOID: RFI
12 06/04/22 14:28

Subject Name

RFI Test

Subject I.D.

Operator Name, I.D.

BC Woods, 782

Location

AS IV Serial no: 111774
Version no: 532B

TEST RECORD 00329

Temp	Date	Time	s/
		210L	

Air Blank:
06/04/22 14:32 .000
Subject Test: Auto
24 06/04/22 14:32 .000

Subject Name

Blank Test

Subject I.D.

Operator Name, I.D.

B.C. Woods, 782

Location