

RECEIVED

By Tracy Crews at 2:51 pm, Aug 08, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN **111776** NAME OF AGENCY **Kansas City MO PD** DATE OF INSPECTION **07/23/2022**

LOCATION OF INSTRUMENT (STREET AND CITY) **9701 Marion Park Drive, Kansas City MO 64137** TIME OF INSPECTION **2019**

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Intoximeters LOT # AG0132803 EXP. DATE 11/24/2023

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE

0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 **.081** TEST 2 **.080** TEST 3 **.080**

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS **0** (0-.04) **0** (.05-.09) **0** (.10-.14) **0** (.15-.19) **0** (OVER .19) **0**

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument meets all DOHSS Standards and Guidelines.

INSPECTING OFFICER

SIGNATURE

PRINT NAME **Pony Carls**

TYPE II PERMIT NUMBER/EXPIRATION DATE **220081 03/16/2024**

TELEPHONE NUMBER **() 816-482-8196**

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS 10 Serial no: 111776
Version no: 532B

TEST RECORD 00320

Temp Date Time 210L
s/

Air Blank:

07/23/22 20:19 .000

Calibration Check:
27 07/23/22 20:19 .081

Subject Name

Test #1

Subject I.D.

Carlsde 4541

Operator Name, I.D.

SPD

Location

AS 10 Serial no: 111776
Version no: 532B

TEST RECORD 00321

Temp Date Time 210L
s/

Air Blank:

07/23/22 20:20 .000

Calibration Check:
27 07/23/22 20:20 .080

Subject Name

Test #2

Subject I.D.

Carlsde 4541

Operator Name, I.D.

SPD

Location

AS 10 Serial no: 111776
Version no: 532B

TEST RECORD 00322

Temp Date Time 210L
s/

Air Blank:

07/23/22 20:22 .000

Calibration Check:
28 07/23/22 20:22 .090

Subject Name

Test #3

Subject I.D.

Carlsde 4541

Operator Name, I.D.

SPD

Location

Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, MO 63103
Ph: (314) 593-3100
Fax: (314) 593-7328

Certificate of Analysis

Test Date: 29-Nov-2021



Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Lot # AG132803 Model 108

Exp Date 24-Nov-2023 Cyl. Type 108 Certified Concentration 0.082 ± 0.002 BrAC (223 ppm)
Component Ethanol Nitrogen

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	206.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm
CRM Serial No. <th>Concentration</th> <td>CRM Serial No. <th>Concentration</th> </td>	Concentration	CRM Serial No. <th>Concentration</th>	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Specific Gravity (20°C) Certified
Reference gas (primary certification of analysis)
Ethanol, 99.9% (v/v) (Lab)
Date: 11/13/2021 13:42

Approved for Release: Rod Marsala
Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

PERMIT TYPE II

COREY CARLISLE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/16/2022 NUMBER 22M081
EXPIRES 3/16/2024
MO 580-0771 (8-19)

Laura G. Noy
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nicholas
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAG- (RS-10)

STATE OF MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM
INSTRUMENT OPERATOR CARD

The owner certifies that the operator is qualified to operate the instrument in accordance with the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

Operator: COREY CARLISLE, COREY
Operator No. 22081
Date Issued 3/16/2022 Date Expires 3/16/2024