



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 7:52 am, Oct 24, 2022

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111781	NAME OF AGENCY Warsaw Police Department	DATE OF INSPECTION 10/23/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 181 w. Harrison Street, Warsaw Mo		TIME OF INSPECTION 10:13 pm

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER Guth Laboratories LOT # 21380 EXP. DATE 09/13/2023
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.01 SIM. SN MP6529 SIM. NIST EXP DATE 05/05/2023
- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .097%	TEST 2 ➡ .098%	TEST 3 ➡ .096%
----------------	----------------	----------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	1	(0-.04)	1	(.05-.09)		(.10-.14)		(.15-.19)	1	(OVER .19)	1
----------	---	---------	---	-----------	--	-----------	--	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument operating with established DOH limits.

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Tony Helms
TYPE II PERMIT NUMBER EXPIRATION DATE 220130 05/11/2024	TELEPHONE NUMBER

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111781
Version no: 532B

TEST RECORD 00075

Temp Date Time ^{a/} 210L

Air Blank:
10/23/22 22:13 .000
Calibration Check:
26 10/23/22 22:13 .097

Subject Name

Test #1

Subject I.D.

Test #1

Operator Name, I.D.

T. Helms 220130

Location

181 W Harrison

Warsaw MO

AS IV Serial no: 111781
Version no: 532B

TEST RECORD 00076

Temp Date Time ^{a/} 210L

Air Blank:
10/23/22 22:15 .000
Calibration Check:
26 10/23/22 22:15 .098

Subject Name

Test #2

Subject I.D.

Test #2

Operator Name, I.D.

T. Helms 220130

Location

181 W Harrison

Warsaw MO

AS IV Serial no: 111781
Version no: 532B

TEST RECORD 00077

Temp Date Time ^{a/} 210L

Air Blank:
10/23/22 22:17 .000
Calibration Check:
27 10/23/22 22:17 .096

Subject Name

Test #3

Subject I.D.

Test #3

Operator Name, I.D.

T. Helms 220130

Location

181 W Harrison

Warsaw MO

AS IV Serial no: 111781
Version no: 532B

TEST RECORD 00078

Temp Date Time ^{a/} 210L

VOID: RFI
12 10/23/22 22:18

Subject Name

RFI Test

Subject I.D.

RFI Test

Operator Name, I.D.

T. Helms 220130

Location

181 W Harrison

Warsaw MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 21380 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on September 15, 2021, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1214% (w/vol) ethyl alcohol. The expiration date for this lot number is September 13, 2023 at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L $\pm 3\%$.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

*Testing was conducted using Cerilliant Reference Standard lot number FN03052002 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.*



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ANTHONY HELMS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/11/2022

Faura Q. Day
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 220130

Dave S. Nickelson

EXPIRES 5/11/2024

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES
LAB-4 (PS-10)

HIC 555 0774 (2.13)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HELMS, ANTHONY
Permit No 220130
Date Issued 5/11/2022 Date Expires 5/11/2024

