

AS IV Serial no: 113778
Version no: 532C

TEST RECORD 00332

Temp Date Time ^{g/} 210L

Air Blank:
10/05/22 17:19 .000
Calibration Check:
21 10/05/22 17:19 .102

Subject Name

Test #1

Subject I.D.

N/A

Operator Name, I.D.

P.O. TORRE 4553

Location

MODOT TMC

AS IV Serial no: 113778
Version no: 532C

TEST RECORD 00333

Temp Date Time ^{g/} 210L

Air Blank:
10/05/22 17:21 .000
Calibration Check:
21 10/05/22 17:21 .102

Subject Name

Test #2

Subject I.D.

N/A

Operator Name, I.D.

P.O. TORRE 4553

Location

MODOT TMC

AS IV Serial no: 113778
Version no: 532C

TEST RECORD 00334

Temp Date Time ^{g/} 210L

Air Blank:
10/05/22 17:22 .000
Calibration Check:
21 10/05/22 17:22 .102

Subject Name

Test #3

Subject I.D.

N/A

Operator Name, I.D.

P.O. TORRE 4553

Location

MODOT TMC

AS IV Serial no: 113778
Version no: 532C

TEST RECORD 00335

Temp Date Time ^{g/} 210L

VOID: RFI
12 10/05/22 17:24

Subject Name

RFI

Subject I.D.

N/A

Operator Name, I.D.

P.O. TORRE 4553

Location

MODOT TMC



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 19-Feb-2022

Lot # AG204801 **Model** 108

Exp Date 17-Feb-2024	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:02.22.2022 19:37

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ROBERT TOSIE

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/24/2022

NUMBER 220173

EXPIRES 6/24/2024

Mike Massman

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES