



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 STATE PUBLIC HEALTH LABORATORY  
 BREATH ALCOHOL PROGRAM

**INTOX EC/IR II MAINTENANCE REPORT**

REPORT #3

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days). Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service. Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX EC/IR II SN 12822	NAME OF AGENCY Harrisonville Police	DATE OF INSPECTION 09/12/2022
----------------------------	----------------------------------------	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 205 N. Lexington Harrisonville	TIME OF INSPECTION 03:10 CDT
----------------------------------------------------------------------------	---------------------------------

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

**DIAGNOSTIC RECORD**

<input checked="" type="checkbox"/> BLANK CHECK	<input checked="" type="checkbox"/> CO2 CHECK
<input checked="" type="checkbox"/> FC 1 TEMP	<input checked="" type="checkbox"/> FLOW CHECK
<input checked="" type="checkbox"/> SRC TEMP	<input checked="" type="checkbox"/> FCB CHECK
<input checked="" type="checkbox"/> DET TEMP	<input checked="" type="checkbox"/> CRC COMP CHECK
<input checked="" type="checkbox"/> BT TEMP	<input checked="" type="checkbox"/> CRC CAL CHECK
<input checked="" type="checkbox"/> STD 2 TEMP	<input checked="" type="checkbox"/> PRINT TEST
<input checked="" type="checkbox"/> ETH CHECK	

**BREATH ANALYZER ACCURACY STANDARDS**

<input type="checkbox"/> SIMULATOR SOLUTION	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER Intoximeters	LOT# AG106801 EXP. DATE 03/09/2023
<input type="checkbox"/> SIMULATOR TEMP (34°C ±0.2°C)	SIM. SN SIM. NIST EXP DATE

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used.

<input type="checkbox"/> 0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
<input checked="" type="checkbox"/> 0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE
<input type="checkbox"/> 0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 0.079 g/210L	TEST 2 0.078 g/210L	TEST 3 0.078 g/210L
---------------------	---------------------	---------------------

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:**

REFUSALS 0	0-.04 0	.05-.09 1	.10-.14 1	.15-.19 1	OVER .19 2
------------	---------	-----------	-----------	-----------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY).

N/A

<b>INSPECTING OFFICER</b>	
SIGNATURE 	PRINT FULL NAME R STARK
TYPE II PERMIT NUMBER 210168	TELEPHONE NUMBER ( 816 ) 380-8940
EXPIRATION DATE 08/04/2023	

**RETURN COMPLETED REPORT TO THE:**  
 Breath Alcohol Program, Missouri Department of Health and Senior Services,  
 by mail, fax, or e-mail



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

Customer Name  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

Test Date: 9-Mar-2021

Lot # AG106801 Model 108cacc

<u>Exp. Date</u> 9-Mar-2023	<u>Cyl. Type</u> 108	<u>Component</u> Ethanol Nitrogen	<u>Certified Concentration</u> 0.080 ± 0.002 BrAC (218 ppm) Balance
--------------------------------	-------------------------	-----------------------------------------	---------------------------------------------------------------------------

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

  

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	253.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Approved for Release: \_\_\_\_\_



ISO 17025:2005 A2LA accredited. Certificate Number 3082.06  
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

2

PERMIT  
 TYPE II

RYAN STARK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/4/2021

NUMBER 210168

EXPIRES 8/4/2023

MO 580:0771 (6-10)

*Laura Q. Way*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Robert K. Kuehl*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB 4, (R6-10)

STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator STARK, RYAN  
 Permit No 210168  
 Date Issued 8/4/2021 Date Expires 8/4/2023



# University of Central Missouri

## Missouri Safety Center - Institute for Public Safety

awards

**RYAN A. STARK**

Certificate of Completion for 8.0 hours of instruction in

### TYPE - II LAB INTOX EC/IR II

The POST approved program for:

No. of CEHs	3.0	Technical Studies
Core Curriculum Area		
No. of CEHs	5.0	Skill Development
Core Curriculum Area		
No. of CEHs	0.0	Interpersonal Perspectives
Core Curriculum Area		
No. of CEHs	0.0	Legal Studies
Core Curriculum Area		

Joanne Kurt-Hilditch, Ph.D.

Senior Director

Matthew Bond

Program Manager

**MISSOURI**

Safety Center

A Missouri Post  
Approved Provider

Given in  
Warrensburg, Missouri  
on July 26, 2021

HARRISONVILLE POLICE DEPT.