



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

**INTOX DMT MAINTENANCE REPORT**

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN <b>500169</b>	NAME OF AGENCY <b>Missouri State Highway Patrol</b>	DATE OF INSPECTION <b>04/02/2022</b>
LOCATION OF INSTRUMENT (STREET AND CITY) <b>Brookfield PD, 116 W. Brooks, Brookfield</b>		TIME OF INSPECTION <b>20:42:33</b>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

**DIAGNOSTIC RECORD**

DATE AND TIME <u>04/02/2022 20:42:36</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.7°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>46.2°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

**BREATH ANALYZER ACCURACY STANDARDS**

SIMULATOR STANDARD       COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER INTOXIMETERS      LOT # AG011501      EXP. DATE 04/24/2022

SIMULATOR TEMP (34°C ± 0.2°C) \_\_\_\_\_      SIM. SN \_\_\_\_\_      SIM. NIST EXP DATE \_\_\_\_\_

**CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)**  
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: <b>0.099</b>	TEST 2: <b>0.099</b>	TEST 3: <b>0.098</b>
----------------------	----------------------	----------------------

**PERFORM R.F.I. TEST**

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 0	0-.04: 0	.05-.09: 0	.10-.14: 0	.15-.19: 0	OVER .19: 0
-------------	----------	------------	------------	------------	-------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

---



---



---




---



---

**INSPECTING OFFICER**

SIGNATURE       PRINT FULL NAME **LEE C EWIGMAN**

TYPE II PERMIT NUMBER **220039**      EXPIRATION DATE **02/10/2024**      TELEPHONE NUMBER **660-385-2132**

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, Missouri Department of Health and Senior Services by mail, fax, or email



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 28-Apr-2020

**Lot # AG011501 Model 108cacc**

<b><u>Exp. Date</u></b> 24-Apr-2022	<b><u>Cyl. Type</u></b> 108	<b><u>Component</u></b> Ethanol Nitrogen	<b><u>Certified Concentration</u></b> 0.100 ± 2% BrAC (260 ppm) Balance
--	--------------------------------	--	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

<table border="0"> <tr><td><b><u>RGM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>EB0010581</td><td>392.1 ppm</td></tr> <tr><td>EB0010570</td><td>259.8 ppm</td></tr> <tr><td>EB0010285</td><td>208.0 ppm</td></tr> <tr><td>EB0010561</td><td>103.6 ppm</td></tr> <tr><td>EB0010681</td><td>52.12 ppm</td></tr> </table>	<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>	EB0010581	392.1 ppm	EB0010570	259.8 ppm	EB0010285	208.0 ppm	EB0010561	103.6 ppm	EB0010681	52.12 ppm	<table border="0"> <tr><td><b><u>RGM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>EB0010603</td><td>393.0 ppm</td></tr> <tr><td>EB0010559</td><td>258.2 ppm</td></tr> <tr><td>EB0010595</td><td>208.3 ppm</td></tr> <tr><td>EB0010562</td><td>104.2 ppm</td></tr> <tr><td>EB0010579</td><td>52.81 ppm</td></tr> </table>	<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>	EB0010603	393.0 ppm	EB0010559	258.2 ppm	EB0010595	208.3 ppm	EB0010562	104.2 ppm	EB0010579	52.81 ppm
<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>																								
EB0010581	392.1 ppm																								
EB0010570	259.8 ppm																								
EB0010285	208.0 ppm																								
EB0010561	103.6 ppm																								
EB0010681	52.12 ppm																								
<b><u>RGM Serial No.</u></b>	<b><u>Concentration</u></b>																								
EB0010603	393.0 ppm																								
EB0010559	258.2 ppm																								
EB0010595	208.3 ppm																								
EB0010562	104.2 ppm																								
EB0010579	52.81 ppm																								
<table border="0"> <tr><td><b><u>CRM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>CC434668</td><td>800.0 ppm</td></tr> <tr><td>CC234503</td><td>253.0 ppm</td></tr> </table>	<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>	CC434668	800.0 ppm	CC234503	253.0 ppm	<table border="0"> <tr><td><b><u>CRM Serial No.</u></b></td><td><b><u>Concentration</u></b></td></tr> <tr><td>0056649</td><td>390.1 ppm</td></tr> <tr><td>0056662</td><td>150.2 ppm</td></tr> </table>	<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>	0056649	390.1 ppm	0056662	150.2 ppm												
<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>																								
CC434668	800.0 ppm																								
CC234503	253.0 ppm																								
<b><u>CRM Serial No.</u></b>	<b><u>Concentration</u></b>																								
0056649	390.1 ppm																								
0056662	150.2 ppm																								

**Analytical Method:** NDIR

Digitally signed by Quality Control  
 Date: 2020.04.28 18:14:39 -05:00  
 Reason: Dry gas standard certification of analysis  
 Location: Airgas USA LLC (Lab)

Approved for Release: \_\_\_\_\_

*Rod Marsala*  
 \_\_\_\_\_  
 Rod Marsala

**ISO 17025:2005 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**LEE C. EWIGMAN**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/10/2022

NUMBER 220039

EXPIRES 2/10/2024

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

, acting director  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator EWIGMAN, LEE  
Permit No 220039  
Date Issued 2/10/2022 Date Expires 2/10/2024

