

RECEIVED

By Tracy Crews at 7:52 am, Apr 28, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
BREATH ALCOHOL PROGRAM

INTOX DMT MAINTENANCE REPORT

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

INTOX DMT SN 500275	NAME OF AGENCY Platte County Sheriff's Office	DATE OF INSPECTION 04/24/2022
LOCATION OF INSTRUMENT (STREET AND CITY) 415 Third Street, Platte City, MO 64079		TIME OF INSPECTION 03:50:49

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC RECORD	
DATE AND TIME <u>04/24/2022 03:50:53</u>	<input checked="" type="checkbox"/> DETECTOR
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTER 1
<input checked="" type="checkbox"/> SAMPLE CHAMBER <u>48.7°C</u>	<input checked="" type="checkbox"/> FILTER 2
<input checked="" type="checkbox"/> BREATH TUBE <u>38.2°C</u>	<input checked="" type="checkbox"/> FILTER 3
<input checked="" type="checkbox"/> PUMP	<input checked="" type="checkbox"/> INTERNAL STANDARD

BREATH ANALYZER ACCURACY STANDARDS

<input type="checkbox"/> SIMULATOR STANDARD	<input checked="" type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>INTOXIMETERS</u>	LOT # <u>AG026606</u> EXP. DATE <u>09/22/2022</u>
<input type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) _____	SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used.

0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE

0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE

0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1: 0.099	TEST 2: 0.098	TEST 3: 0.099
---------------	---------------	---------------

PERFORM R.F.I. TEST

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:

REFUSALS: 1	0-.04: 1	.05-.09: 2	.10-.14: 10	.15-.19: 5	OVER .19: 4
-------------	----------	------------	-------------	------------	-------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS. (USE OTHER SIDE IF NECESSARY)

INSPECTING OFFICER	
SIGNATURE <i>Caleb Jeffries</i>	PRINT FULL NAME CALEB M JEFFRIES
TYPE II PERMIT NUMBER 200248	EXPIRATION DATE 09/22/2022
	TELEPHONE NUMBER 816-858-1803

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, Missouri Department of Health and Senior Services by mail, fax, or email



Airgas USA LLC (LAB)
 3600 Bernard Street
 St. Louis, Mo. 63108
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63148

Test Date: 23-Sep-2020

Lot # AG026606 Model 108cadd

Exp. Date 22-Sep-2022	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0,100 ± 2% BrAC (280 ppm) Balance
---------------------------------	-------------------------	---	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<table border="0"> <tr><td>RGM Serial No.</td><td>Concentration</td></tr> <tr><td>EB0010581</td><td>392.1 ppm</td></tr> <tr><td>EB0010570</td><td>258.8 ppm</td></tr> <tr><td>EB0010285</td><td>208.0 ppm</td></tr> <tr><td>EB0010581</td><td>103.6 ppm</td></tr> <tr><td>EB0010581</td><td>62.12 ppm</td></tr> </table>	RGM Serial No.	Concentration	EB0010581	392.1 ppm	EB0010570	258.8 ppm	EB0010285	208.0 ppm	EB0010581	103.6 ppm	EB0010581	62.12 ppm	<table border="0"> <tr><td>RGM Serial No.</td><td>Concentration</td></tr> <tr><td>EB0010603</td><td>393.0 ppm</td></tr> <tr><td>EB0010559</td><td>258.2 ppm</td></tr> <tr><td>EB0010595</td><td>208.3 ppm</td></tr> <tr><td>EB0010582</td><td>104.2 ppm</td></tr> <tr><td>EB0010579</td><td>62.61 ppm</td></tr> </table>	RGM Serial No.	Concentration	EB0010603	393.0 ppm	EB0010559	258.2 ppm	EB0010595	208.3 ppm	EB0010582	104.2 ppm	EB0010579	62.61 ppm
RGM Serial No.	Concentration																								
EB0010581	392.1 ppm																								
EB0010570	258.8 ppm																								
EB0010285	208.0 ppm																								
EB0010581	103.6 ppm																								
EB0010581	62.12 ppm																								
RGM Serial No.	Concentration																								
EB0010603	393.0 ppm																								
EB0010559	258.2 ppm																								
EB0010595	208.3 ppm																								
EB0010582	104.2 ppm																								
EB0010579	62.61 ppm																								
<table border="0"> <tr><td>CRM Serial No.</td><td>Concentration</td></tr> <tr><td>CC434668</td><td>800.0 ppm</td></tr> <tr><td>GC234503</td><td>253.0 ppm</td></tr> </table>	CRM Serial No.	Concentration	CC434668	800.0 ppm	GC234503	253.0 ppm	<table border="0"> <tr><td>CRM Serial No.</td><td>Concentration</td></tr> <tr><td>0056649</td><td>390.1 ppm</td></tr> <tr><td>0056662</td><td>160.2 ppm</td></tr> </table>	CRM Serial No.	Concentration	0056649	390.1 ppm	0056662	160.2 ppm												
CRM Serial No.	Concentration																								
CC434668	800.0 ppm																								
GC234503	253.0 ppm																								
CRM Serial No.	Concentration																								
0056649	390.1 ppm																								
0056662	160.2 ppm																								

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2020.09.23 16:39:18 -0500
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: Rod Marsala
 Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CALEB M JEFFRIES

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 9/22/2020

NUMBER 200248

EXPIRES 9/22/2022

MS-300077 (1) 10

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MS-300077 (1) 10

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The instrument operator is authorized to operate all instrument breath alcohol analyzers throughout the determination of the alcoholic content of breath from persons in Missouri.

Operator: **JEFFRIES, CALEB**
Permit No: **200248**
Date Issued: **9/22/2020** Date Expires: **9/22/2022**