



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
CMi INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 2:26 pm, Feb 01, 2022

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005827	LOCATION OF INSTRUMENT ORONOGO POLICE DEPT	DATE OF INSPECTION 02/01/2022	TIME OF INSPECTION 10:56
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	10:58	DRY	AG026705	09/23/2022
Cal Check	0.081	10:58	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	10:59	N/A	N/A	N/A
Cal Check	0.081	10:59	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	10:59	0.080	INTOXIMETERS, INC	
Cal Check	0.081	11:00	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	11:00	0.081		
			CALIBRATION CHECK RESULT 2		
			0.081		
			CALIBRATION CHECK RESULT 3		
			0.081		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.2%	0.000	

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		-----		
RAM Test	Pass		Air Blank	0.000	11:01
EEPROM Checksum Test	Pass		Subject Test	RFI*	11:01
Real Time Clock Test	Pass		Air Blank	0.000	11:01
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	00-04	05-09	10-14	15-19	OVER 19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER			
SIGNATURE 	PRINT NAME CHRISTOPHER SHONK		
TYPE II PERMIT NUMBER 210136	EXPIRATION DATE 06/30/2023	TELEPHONE NUMBER 4176731911	



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 28-Sep-2020

Lot # AG026705 Model 108cacc

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
23-Sep-2022	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (208 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:


<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 Date: 2020.09.30 14:49:51 -05 00
 Reason: Dry gas standard certification of analysis
 Location: Airgas USA LLC (Lab)

Approved for Release: _____



Rod Marsala

ISO 17025:2005 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CHRISTOPHER SHONK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT, INTOXILYZER 8000, ALCO-SENSOR IV
 W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE _____

NUMBER 6/30/2021

EXPIRES 210136

MO 580-0771 (6-1) 6/30/2023

Laura Q. Nay RATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

 STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SHONK, CHRISTOPHER
 Permit No 210136
 Date Issued 6/30/2021 Date Expires 6/30/2023

