



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
**CMI INTOXILYZER 8000 MAINTENANCE REPORT**

**RECEIVED**

By Tracy Crews at 8:30 am, Aug 09, 2022

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER <b>80-005827</b>	LOCATION OF INSTRUMENT <b>ORONOGO POLICE DEPT</b>	DATE OF INSPECTION <b>08/03/2022</b>	TIME OF INSPECTION <b>21:45</b>
--	--	---	------------------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE <b>DRY</b>	STANDARD LOT # <b>AG026705</b>	STANDARD EXPIRATION DATE <b>09/23/2022</b>
-----	-----	-----	SIM TEMPERATURE <b>N/A</b>	SIM SERIAL NUMBER <b>N/A</b>	SIM CERTIFICATE EXPIRATION <b>N/A</b>
Air Blank	0.000	21:47	STANDARD VALUE <b>0.080</b>	STANDARD SUPPLIER <b>INTOXIMETERS, INC</b>	
Cal Check	0.081	21:47	CALIBRATION CHECK RESULT 1 <b>0.081</b>		
Air Blank	0.000	21:48	CALIBRATION CHECK RESULT 2 <b>0.081</b>		
Cal Check	0.081	21:48	CALIBRATION CHECK RESULT 3 <b>0.081</b>		
Air Blank	0.000	21:49	MAXIMUM DEVIATION (MUST BE WITHIN 5%) <b>1.2%</b>		
Cal Check	0.081	21:49	SPREAD (MUST BE .005 OR LESS) <b>0.000</b>		
Air Blank	0.000	21:49			

**Pass**

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		-----	-----	-----
EEPROM Checksum Test	Pass		Air Blank	0.000	21:50
Real Time Clock Test	Pass		Subject Test	RFI*	21:50
DSP Test	Pass		Air Blank	0.000	21:51
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass				
Temperature Regulation Test	Pass				

**Pass**

**Pass**

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT					
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

<b>INSPECTING OFFICER</b>		PRINT NAME	
SIGNATURE 		CHRISTOPHER SHONK	
TYPE II PERMIT NUMBER <b>210136</b>	EXPIRATION DATE <b>06/30/2023</b>	TELEPHONE NUMBER <b>4176731911</b>	







STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**CHRISTOPHER SHONK**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT, INTOXILYZER 8000, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/30/2021

NUMBER 210136

EXPIRES 6/30/2023

MO 585-0771 (6-10)

*Laura A. Day*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Robert Knudsen*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (P5-12)

STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator SHONK, CHRISTOPHER  
Permit No 210136  
Date issued 6/30/2021 Date Expires 6/30/2023