



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED

By Tracy Crews at 1:34 pm, Nov 07, 2022

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005827	LOCATION OF INSTRUMENT ORONOGO POLICE DEPT	DATE OF INSPECTION 11/04/2022	TIME OF INSPECTION 18:35
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	18:37	DRY	34221080A2	01/05/2024
Cal Check	0.080	18:37	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	18:38	N/A	N/A	N/A
Cal Check	0.079	18:38	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	18:39	0.080	ILMO	
Cal Check	0.078	18:39	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	18:39	0.080		
Cal Check	0.078	18:39	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	18:39	0.079		
Cal Check	0.078	18:39	CALIBRATION CHECK RESULT 3		
Air Blank	0.000	18:39	0.078		
Pass			MAXIMUM DEVIATION (MUST BE WITHIN 5%)		SPREAD (MUST BE .005 OR LESS)
			2.5%		0.002

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
			Test	g/210L	Time
Voltage/Current Test	Pass		-----	-----	-----
RAM Test	Pass		Air Blank	0.000	18:40
EEPROM Checksum Test	Pass		Subject Test	RFI*	18:40
Real Time Clock Test	Pass		Air Blank	0.000	18:41
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				
Pass			Pass		

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME CHRISTOPHER SHONK	
TYPE II PERMIT NUMBER 210136		EXPIRATION DATE 06/30/2023	TELEPHONE NUMBER 4176731911



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 14513
Part #: BAC34L080T
Cylinder Size: 34L
Lot Number: 34221080A2
Expiration: 1/5/2024

360-08KT

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 34 Liters @ 500 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	288 ppm	+/- 0.002 BAC (G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

*Traceable to:
Certified Reference Material - 262.4 µmol/mol
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

12-17-2021
Issuance Date



The calibration results on this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
CHRISTOPHER SHONK

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOX DMT, INTOXILYZER 8000, ALCO-SENSOR IV W/PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/30/2021

NUMBER 210136

EXPIRES 6/30/2023

MO 585-0771 (6-10)

Laura A. Day

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Robert Knudsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (P5-12)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.



Operator SHONK, CHRISTOPHER
Permit No 210136
Date issued 6/30/2021 Date Expires 6/30/2023