



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
CMI INTOXILYZER 8000 MAINTENANCE REPORT

**RECEIVED**

By Tracy Crews at 8:22 am, Jan 03, 2023

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005827	LOCATION OF INSTRUMENT ORONOGO POLICE DEPT	DATE OF INSPECTION 12/31/2022	TIME OF INSPECTION 04:48
---------------------------------------	---	----------------------------------	-----------------------------

**CALIBRATION CHECK RESULTS**

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	04:49
Cal Check	0.081	04:49
Air Blank	0.000	04:50
Cal Check	0.080	04:50
Air Blank	0.000	04:51
Cal Check	0.080	04:51
Air Blank	0.000	04:52

Pass

**CALIBRATION CHECK SUMMARY**

STANDARD TYPE DRY	STANDARD LOT # 34221080A2	STANDARD EXPIRATION DATE 01/05/2024
SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
STANDARD VALUE 0.080	STANDARD SUPPLIER ILMO	
CALIBRATION CHECK RESULT 1 0.081		
CALIBRATION CHECK RESULT 2 0.080		
CALIBRATION CHECK RESULT 3 0.080		
MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%	SPREAD (MUST BE .005 OR LESS) 0.001	

**DIAGNOSTIC TEST RESULTS**

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

**RFI TEST RESULTS**

Test	g/210L	Time
-----	-----	-----
Air Blank	0.000	04:52
Subject Test	RFI*	04:52
Air Blank	0.000	04:53
*RFI Detect		

Pass

**NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT**

REFUSALS	00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	0	0	0	0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE 	PRINT NAME CHRISTOPHER SHONK
---------------	---------------------------------

TYPE B PERMIT NUMBER 210136	EXPIRATION DATE 06/30/2023	TELEPHONE NUMBER 4176731911
--------------------------------	-------------------------------	--------------------------------



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

### Certificate of Analysis

Certificate ID: 14513  
Part #: BAC34L080T  
Cylinder Size: 34L  
Lot Number: 34221080A2  
Expiration: 1/5/2024

*360-08KT*

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 34 Liters @ 500 psig 70°F (21°C)

Component	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:
Ethanol	288 ppm	+/- 0.002 BAC (G/210L) [5.2 ppm]	NDIR
Nitrogen	Balance		

\*Traceable to:  
Certified Reference Material - 262.4 µmol/mol  
Ethanol in Nitrogen - Serial No. GN0015026 Lot No. 050319E11

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

*[Signature]*  
Specialty Gas Lab Tech  
*[Signature]*

12-17-2021  
Issuance Date



The calibration results on this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items contained on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

2

**PERMIT**  
**TYPE II**  
**CHRISTOPHER SHONK**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT, INTOXILYZER 8000, ALCO-SENSOR IV W/PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/30/2021

NUMBER 210136

EXPIRES 6/30/2023

MO 583-0771 (6-10)

*Laura A. Day*  
\_\_\_\_\_  
DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Rob Knudt*  
\_\_\_\_\_  
DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (RS-12)



**STATE OF MISSOURI**  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*



Operator SHONK, CHRISTOPHER  
Permit No 210136  
Date issued 6/30/2021 Date Expires 6/30/2023