



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

RECEIVED
07/14/2022 11:22 am, JPL

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005835	LOCATION OF INSTRUMENT IPD DETENTION	DATE OF INSPECTION 07/14/2022	TIME OF INSPECTION 16:25
---------------------------------------	---	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # AG117305	STANDARD EXPIRATION DATE 06/22/2023
Air Blank	0.000	16:27	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.080	16:27	STANDARD VALUE 0.080	STANDARD SUPPLIER INTOXIMETERS, INC	
Air Blank	0.000	16:27	CALIBRATION CHECK RESULT 1 0.080		
Cal Check	0.079	16:28	CALIBRATION CHECK RESULT 2 0.079		
Air Blank	0.000	16:28	CALIBRATION CHECK RESULT 3 0.079		
Cal Check	0.079	16:29	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 1.2%		
Air Blank	0.000	16:29	SPREAD (MUST BE .005 OR LESS) 0.001		

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	16:30
EEPROM Checksum Test	Pass		Subject Test	RFI*	16:30
Real Time Clock Test	Pass		Air Blank	0.000	16:30
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass		<h1>Pass</h1>		
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT					
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
5	1	1	2	0	4

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSTRUMENT OPERATED WITHIN STATE STANDARDS

INSPECTING OFFICER		
SIGNATURE <i>Ron Baltzer #1460</i>	PRINT NAME RON BALTZER	
TYPE II PERMIT NUMBER 210144	EXPIRATION DATE 07/16/2023	TELEPHONE NUMBER 816-325-7300



Airgas USA LLC (LAB)

3500 Bernard Street

St. Louis, Mo. 63103

Ph: (314) 533-3100

Fax: (314) 533-7328

Certificate of Analysis

Customer Name

Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 23-Jun-2021

Lot # AG117305 Model 108cadd

<u>Exp. Date</u>	<u>Cyl. Type</u>	<u>Component</u>	<u>Certified Concentration</u>
22-Jun-2023	108	Ethanol Nitrogen	0.080 ± 0.002 BrAC (208 ppm) Balance

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

<u>RGM Serial No.</u>	<u>Concentration</u>	<u>RGM Serial No.</u>	<u>Concentration</u>
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

<u>CRM Serial No.</u>	<u>Concentration</u>	<u>CRM Serial No.</u>	<u>Concentration</u>
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
Date: 2021.06.24 18:10:42 -05:00
Reason: Dry gas standard certification of analysis
Location: Airgas USA LLC (Lab)

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
RON BALTZER

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 7/16/2021

Laura Q. Nacy

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

NUMBER 210144

Ron Kuehl

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

EXPIRES 7/16/2023

 **STATE OF MISSOURI**
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **BALTZER, RON**
 Permit No **210144**
 Date Issued **7/16/2021** Date Expires **7/16/2023**

