

RECEIVED

By Tracy Crews at 3:04 pm, Mar 15, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005843	LOCATION OF INSTRUMENT KANSAS CITY P.D.	DATE OF INSPECTION 03/05/2022	TIME OF INSPECTION 21:27
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
-----	-----	-----	DRY	14020080A2	07/05/2022
Air Blank	0.000	21:29	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Cal Check	0.082	21:29	N/A	N/A	N/A
Air Blank	0.000	21:30	STANDARD VALUE	STANDARD SUPPLIER	
Cal Check	0.080	21:30	0.080	CMI	
Air Blank	0.000	21:31	CALIBRATION CHECK RESULT 1		
Cal Check	0.080	21:31	0.082		
Air Blank	0.000	21:31	CALIBRATION CHECK RESULT 2		
Cal Check	0.080	21:31	0.080		
Air Blank	0.000	21:31	CALIBRATION CHECK RESULT 3		
Pass			MAXIMUM DEVIATION (MUST BE WITHIN 5%)		SPREAD (MUST BE .005 OR LESS)
			2.5%		0.002

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass	Test	g/210L	Time	
RAM Test	Pass	-----	-----	-----	
EEPROM Checksum Test	Pass	Air Blank	RFI*	21:32	
Real Time Clock Test	Pass	Air Blank	0.000	21:32	
DSP Test	Pass	*RFI Detect			
Analytical Stability Test	Pass	Pass			
Internal Printer Test	Pass				
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT							
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19		
0	5	2	4	1	6		

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

TESTED AND CERTIFIED

INSPECTING OFFICER			
SIGNATURE 		PRINT NAME INFRANCA, JORDAN	
TYPE II PERMIT NUMBER 210128		EXPIRATION DATE 06/21/2023	TELEPHONE NUMBER 816-382-5897



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

Certificate of Analysis

Certificate ID: 13021
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14020080A2
Expiration: 7/5/2022

0.80 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component	Reported Concentration	Analytical Accuracy (U, k=2)	Analytical Method
ethanol	208 ppm	+/-0.032 BAC (g/210L)	NDIR
nitrogen	balance	[5.2 ppm]	

Distributed by:
CMI Inc.
316 East Ninth Street
Owensboro, KY 42303
Phone 866-835-0690
www.alcoholtest.com

Traceable to:
Certified Reference Material - 262.4 µmol/mol
nitrogen in Nitrogen - Serial No. GN0015026 Lot No. 050319E1

Store in dry area, away from sources of heat,
ignition and direct sunlight. Do not allow storage
area to exceed 52 °C (125 °F).

Specialty Gas Lab Tech

Travis Nelson

06-15-2020
Issuance Date



Calibration results within this certificate were obtained using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items listed on this certificate. ILMO Product Company makes no warranty or representation as to the suitability of the use of any information provided for any particular use. This information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service.

ISO/IEC 17025:2017 Accredited Laboratory



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT TYPE II

JORDAN INFRANCA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 308.111 through 308.119 RSMo.

DATE 6/21/2019

NUMBER 290130

EXPIRES 6/21/2021

MO 660-0771 (6-10)

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R0-10)

