

RECEIVED

By Tracy Crews at 7:19 am, Feb 18, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005852	LOCATION OF INSTRUMENT JOPLIN POLICE DEPT	DATE OF INSPECTION 02/11/2022	TIME OF INSPECTION 08:51
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS

Test	g/210L	Time
Air Blank	0.000	08:55
Cal Check	0.102	08:55
Air Blank	0.000	08:56
Cal Check	0.102	08:56
Air Blank	0.000	08:57
Cal Check	0.101	08:57
Air Blank	0.000	08:57

Pass

CALIBRATION CHECK SUMMARY

STANDARD TYPE DRY	STANDARD LOT # AG125707	STANDARD EXPIRATION DATE 09/14/2023
SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
STANDARD VALUE 0.100	STANDARD SUPPLIER INTOXIMETERS	
CALIBRATION CHECK RESULT 1 0.102		
CALIBRATION CHECK RESULT 2 0.102		
CALIBRATION CHECK RESULT 3 0.101		
MAXIMUM DEVIATION (MUST BE WITHIN 5%) 2.0%	SPREAD (MUST BE .005 OR LESS) 0.001	

DIAGNOSTIC TEST RESULTS

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temperature Regulation Test	Pass

Pass

RFI TEST RESULTS

Test	g/210L	Time
Air Blank	0.000	08:58
Subject Test	RFI*	08:58
Air Blank	0.000	08:59
*RFI Detect		

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	0	3	7	1	4

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

NONE

INSPECTING OFFICER

SIGNATURE 	PRINT NAME MORA, NETISHA
---------------	-----------------------------

TYPE II PERMIT NUMBER 210165	EXPIRATION DATE 08/04/2023	TELEPHONE NUMBER 9564011447
---------------------------------	-------------------------------	--------------------------------



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
NETISHA MORA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **8/4/2021**

NUMBER **210165**

EXPIRES **8/4/2023**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MORA, NETISHA
Permit No 210165
Date Issued 8/4/2021 **Date Expires** 8/4/2023