



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005852		LOCATION OF INSTRUMENT JOPLIN POLICE DEPT		DATE OF INSPECTION 06/17/2022		TIME OF INSPECTION 08:56		
CALIBRATION CHECK RESULTS				CALIBRATION CHECK SUMMARY				
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE			
			DRY	AG1 25707	09/14/2023			
			SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION			
			N/A	N/A	N/A			
			STANDARD VALUE	STANDARD SUPPLIER				
			0.100	INTOXIMETERS				
			CALIBRATION CHECK RESULT 1			0.100		
CALIBRATION CHECK RESULT 2			0.100					
CALIBRATION CHECK RESULT 3			0.101					
MAXIMUM DEVIATION (MUST BE WITHIN 5%)			SPREAD (MUST BE .005 OR LESS)					
1.0%			0.001					
DIAGNOSTIC TEST RESULTS				RFI TEST RESULTS				
Voltage/Current Test Pass				Test				
RAM Test Pass				g/210L				
EEPROM Checksum Test Pass				Time				
Real Time Clock Test Pass				Air Blank 0.000 09:03				
DSP Test Pass				Subject Test RFI* 09:03				
Analytical Stability Test Pass				Air Blank 0.000 09:04				
Modem Test Pass				*RFI Detect				
Temperature Regulation Test Pass								
Pass				Pass				
NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT								
REFUSALS	.00-.04	.05-.09	10-.14	.15-.19	OVER .19			
0	0	1	4	2	1			
List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary). NONE								
INSPECTING OFFICER								
SIGNATURE <i>Netisha Mora</i>				PRINT NAME MORA, NETISHA				
TYPE II PERMIT NUMBER 210165		EXPIRATION DATE 08/04/2023		TELEPHONE NUMBER 9564011447				



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 15-Sep-2021

Lot # AG125707 **Model** 55

Exp Date	Cyl. Type	Component	Certified Concentration
14-Sep-2023	55	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Digitally signed by:Quality Control
Reason:Dry gas standard certification of analysis
Location:Airgas USA LLC (Lab)
Date:09.15.2021 19:37

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
NETISHA MORA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/4/2021

NUMBER 210165

EXPIRES 8/4/2023

Laura A. Nagy

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Robt. Knudsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R6-10)

MO 680-0771 (6-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MORA, NETISHA
 Permit No 210165
 Date Issued 8/4/2021 Date Expires 8/4/2023

