

RECEIVED

By Brian Lutmer at 9:22 am, Jul 15, 2022



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005852	LOCATION OF INSTRUMENT JOPLIN POLICE DEPT	DATE OF INSPECTION 07/15/2022	TIME OF INSPECTION 07:15
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	07:18	DRY	AG125707	09/14/2023
Cal Check	0.101	07:18	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	07:18	N/A	N/A	N/A
Cal Check	0.101	07:19	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	07:19	0.100	INTOXIMETERS	
Cal Check	0.101	07:20	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	07:20	0.101		
			CALIBRATION CHECK RESULT 2		
			0.101		
			CALIBRATION CHECK RESULT 3		
			0.101		
			MAXIMUM DEVIATION (MUST BE WITHIN 5%)	SPREAD (MUST BE .005 OR LESS)	
			1.0%	0.000	

Pass

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass				
EEPROM Checksum Test	Pass		Air Blank	0.000	07:21
Real Time Clock Test	Pass		Subject Test	RFI*	07:21
DSP Test	Pass		Air Blank	0.000	07:21
Analytical Stability Test	Pass		*RFI Detect		
Modem Test	Pass				
Temperature Regulation Test	Pass				

Pass

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
0	1	1	5	4	3	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).
NONE

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME MORA, NETISHA	
TYPE II PERMIT NUMBER 210165	EXPIRATION DATE 08/04/2023	TELEPHONE NUMBER 9564011447



Airgas USA LLC (LAB)
3500 Bernard Street
St. Louis, Mo. 63103
Ph: (314) 533-3100
Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
Intoximeters, Inc.
2081 Craig Road
St. Louis, Mo 63146

Test Date: 15-Sep-2021

Lot # AG125707 **Model** 55

Exp Date	Cyl. Type	Component	Certified Concentration
14-Sep-2023	55	Ethanol Nitrogen	0.100 ± 2% BrAC (260 ppm)

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	392.1 ppm	EB0010603	393.0 ppm
EB0010570	259.8 ppm	EB0010559	258.2 ppm
EB0010285	208.0 ppm	EB0010595	208.3 ppm
EB0010561	103.6 ppm	EB0010562	104.2 ppm
EB0010681	52.12 ppm	EB0010579	52.81 ppm

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC434668	800.0 ppm	0056649	390.1 ppm
CC234503	253.0 ppm	0056662	150.2 ppm

Analytical Method: NDIR

Approved for Release: _____

Rod Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
NETISHA MORA

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE **8/4/2021**

NUMBER **210165**

EXPIRES **8/4/2023**

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator MORA, NETISHA
Permit No 210165
Date Issued 8/4/2021 **Date Expires** 8/4/2023